

# Invitation of Quotation


For

Department of Anatomy

At

All India Institute of Medical Sciences, Rishikesh

**Issue Date** : 20 Jan, 2018  
**Inquiry No.** : F.No.32/451/2018(RIS)  
**Last Date of Submission** : 27 Jan, 2018 at 05:00 PM.



## All India Institute of Medical Sciences, Rishikesh

Virbhadrā Road, Rishikesh, Uttarakhand-249203.  
Telephone: 0135-2462913 , [www.aiimsrishikesh.edu.in](http://www.aiimsrishikesh.edu.in)



## Invitation of Quotation for Department Of Anatomy.

Sealed Quotations are hereby invited by the undersigned on behalf of the Director, AIIMS Rishikesh for supply of Items for the Institute as per terms & conditions mentioned below. The filled quotations along with all the required document must reach in the office of the undersigned on or before 27.01.2018, 05.00 PM. The Envelope containing the quotation would please be sealed and super scribed as under:-

### " QUOTATION FOR Department Of Anatomy.

FILE NO. 32/451/2018(RIS) DUE ON 27.01.2017, 05.00 PM"

#### 1. Terms & Conditions:

- A) The quotations received after this deadline & unsealed shall not be entertained under any circumstances whatsoever. In case of postal delay this Institute will not be responsible.
- B) Quotations must be in the enclosed prescribed Performa on the letter head of the firm duly signed by the Proprietor/ Partner/ Director or their authorized representative, In case of signing of quotation by the authorized representative letter of authorization must be attached with the quotation.
- C) Rates must be quoted as per the format specified Taxes extra if any must be written separately.
- D) No overwriting or cutting is permitted in the rate. If found, the quotation shall be summarily rejected.
- E) The rates quoted must be valid for 60 days minimum from the date of opening of the quotation and silence of any tendered on this issue shall be treated as agreed with this condition.
- F) Becoming L1 will not be the criteria for awarding of purchase order unless the rates are reasonable & justified.  
**L1 will be decided on individual item basis.**
- G) RTGS/NEFT details need to be furnished by the supplier with the quotation on the letter head of supplier/firm/agency.
- H) The firm/agency may satisfy the following conditions and attach self-attested copy of the same with the quotation:
  - Firm shall be registered with the Government of Uttarakhand/Central Govt.
  - The firm shall have valid GST/ Sales Tax No. and IT PAN.
  - **The firm should not be black listed by any Govt. Agency/Dept.**
- I) Quotations qualified by such vague and indefinite expressions such as "subject to prior confirmation", "subject to immediate acceptance" etc. will be treated as vague offers and rejected accordingly. Any conditional quotation shall be rejected summarily.
- J) The supplier may be asked to submit a sample of the product(s), which will be evaluated by a technical evaluation committee. The expenditure incurred for demonstrating the items will be borne by the supplier.
- K) **Delivery Period – 15 days from issuing purchase order.**



- L) **Liquidated Damage:** - If the supplier fails to deliver the material on or before the stipulated date, then a penalty at the rate of 0.5 % per week of the total order value shall be levied subject to maximum of 10% of the total order value.
- M) **Payment Terms:** Payment will be made only after satisfactorily delivery, commissioning and inspection of material by the AIIMS Rishikesh.
- N) **Disputes:** -In the event of any dispute or disagreement arising between the Supplier and any other department of AIIMS Rishikesh with regards to the interpretation of "Terms & Conditions" of this inquiry, the same shall be referred to the Director, AIIMS Rishikesh whose decision will be final and binding upon the Supplier.
- O) AIIMS, Rishikesh reserves the right to increase or decrease quantity and / or amount of work. Decision of Quantity of material in the AIIMS, Rishikesh will be final in this regard.
- P) AIIMS, Rishikesh reserves the right to reject any quotation or part or the whole of inviting quotation process without assigning any reason. Decision of the AIIMS, Rishikesh will be final in this regard.

**Administrative Officer**

Encl.: Annexure 1 (Format of Price Bid)



[On the letterhead of firm]

**ANNEXURE - I**  
**PRICE BIDFORM**

To,  
Administrative Officer,  
AIIMS, Rishikesh.

Dear Sir,

1. I/We ..... Submitted the quotation for Enquiry No. "QUOTATION for Department of Anatomy AGAINST THE INQUIRY NO. FILE NO. 32/451/2018(RIS)" DUE ON 27.01.2018, 05.00 PM for Supply of following items at AIIMS Rishikesh".
2. I/We thoroughly examined, understood and accepted terms & conditions given in the enquiry document, failing which my quotation will be rejected out rightly.
3. I/We hereby offer to supply at the following rates.

S. No	Items	Qty.
1.	Single Tube Multi Mode Reader ( Luminometer / Flurometer/ Absorbance Photometer)	01

Date \_\_\_\_\_

Place \_\_\_\_\_

Signature of Authorized Person) :- \_\_\_\_\_

(Name) \_\_\_\_\_

Name of Firm/Company/Agency \_\_\_\_\_

Phone No. \_\_\_\_\_

Email: \_\_\_\_\_



### Microplate Luminometer

- System should have Luminescence, Fluorescence and Visible Absorbance measurement capabilities.
- System should be capable for performing applications such as – Cell health assay, Cell Signaling and metabolism assays, Reporter gene Assays, Nucleic Acid and Protein Quantitation (fluorescent dye based), Elisa Assays, BCA protein Assays, Bradford Protein Assays, Multiplex Assays (Cell Viability, Cytotoxicity, and Apoptosis Assays)
- Compatible with third-party automation control. SiLA Compliant.
- Sample Format : 6, 12, 24, 48, 96 and 384 well plates
- Inbuilt Heater with Temperature Range Ambient +5degc to 45 Deg C
- Inbuilt Shaker with linear or orbital shaking patterns.
- Integrated Tablet PC for instrument operation. Wireless connectivity to wireless network, Data Exportable to USB Flash drive.
- Data exportable to Excel spreadsheet for data analysis.
- System should have Preloaded assays Protocol, and provision for creating customized protocols.
- Integrated Data Analysis Software – Should be able to Label Blank, Known and Unknown Samples, Assign replicates and perform curve fit analysis of data on the instrument.

#### Luminescence Module:

- Detector : Top Reading, Head on Photon Counting Multiplier Tube
- Detection Limit:  $3 \times 10^{-21}$  moles of Luciferase
- Linear Dynamic Range of 9 Logs without the use of filters or manual gain changes.
- Cross Talk: Less than  $3 \times 10^{-5}$  (white 96 well plate)
- System Qualified as DLReady™ and NanoLuc™ Ready

#### Fluorescence Module:

- Detector: PIN –photo Diode
- Light Source: Wavelength Matched LED
- Read Position: Top Reading
- 5 standard excitation and emission filters : Wavelength included - UV (Ex: 365nm, Em: 415- 445nm); Blue (Ex: 475nm, Em: 500 -550nm); Green (Ex: 525nm, Em: 580- 640nm); Red (Ex: 625nm, Em: 660 -720nm); AFC (Ex: 405nm, Em: 495 - 505nm)
- One empty position for customizable excitation and emission wavelength
- Detection Limit : 2fmol Fluorescein /200ul
- Linear Dynamic Range : >6 logs

#### Visible Absorbance Module

- Detector: Top-reading, head-on photon counting photomultiplier tube (PMT)
- Light Source: Xenon flash lamp
- Filter Wheel: 5 absorbance filters Wavelengths 405, 450, 490, 560 and 600nm ,10nm bandwidth
- Detection Limit 0.1 O.D.
- Dynamic Range: 0–4.0 O.D. in the 405–600nm spectral range

Certified with CE/BIS/ISO

**Optional** -Dual Reagent Injector system with dispense volume range of 5-200ul in 1ul increments, compatible with 6-, 12-, 24-, 48- and 96-well plates.

*Qah*

*SFC*  
*vreg*

*BR*

*φ*