

# **Corrigendum**

## **In tender document Tender Enquiry No.24/RT/130/2016-RISH (ADMN)**

Dated: 22-06-2017

As per schedule, Pre- Bid meeting of "Tender for Flexible Fibre-Optic Bronchoscope with Camera & Light Source for Department of Pulmonary Medicine" was held on 12-06-2017 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No. 24/RT/130/2017-RISH (ADMN) are being made.

### **Point srl no. 01 at page no. 20: -**

**For: -** It should have superior image quality with crisp, clear images and true to life colour preferably HD images.

**Read as: -** It should have superior image quality with crisp, clear images and true to life colour.

### **Point srl no 03 at page no. 20: -**

**For: -** Channel diameter should be 2.0mm or more.

**Read as: -** Channel diameter should be atleast 2.8 mm or more.

### **Point srl no 14 at page no. 20: -**

**For:-** Light Source- Preferably Xenon.

**Read as: -** Light Source- Xenon/LED/Halogen, preferably Xenon.

### **Point srl no 01 at page no. 21: -**

**For: -** Compact trolley should be made of stainless steel capable of carrying the weight of all the equipment. Should have multiple shelves to keep accessories mentioned.

**Read as: -** Compact locally supplied trolley should be made of stainless steel capable of carrying the weight of all the equipment. Should have multiple shelves to keep accessories mentioned.

**Note:** Last date of submission of bids in respect of aforesaid tender is hereby extended till **12-08-2017** at 03:00 PM and Technical bid will be opened on same day at 03.30 PM.