

Corrigendum

In tender document Tender Enquiry No.24/RT/132(X)/2017-RISH (ADMN)

Dated: 07-07-2017

As per schedule, Pre- Bid meeting of "Tender for Elbow Continuous Passive Motion Machine for the Department of Orthopedic" was held on 29-06-2017 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No. 24/RT/Orhto//132(X)2017-RISH (ADMN) has been made.

Point srl no. 01 at Page No. 20:

For: - Therapy modes: Mobilization modes for continuous, intermittent and progressive passive movements.

Read as: - Therapy modes: Mobilization modes for continuous, intermittent and passive movements.

Point srl no. 06 at Page No. 20:

For: - Memory/ Programme: 80 pre-set treatment patterns are stored in memory.

Read as: - Deleted.

Point srl no. 07 at Page No. 20:

For: - Open/ Programmes: 20 programmes are free to store your own patterns.

Read as:- Deleted.

Point srl no. 09 at Page No. 20:

For: - Height Adjustment: Electrically, by push button switches.

Read as: - Height Adjustment: mechanically/Electrically, by push button switches.

Note: Last date of submission of bids in respect of aforesaid tender is hereby extended till **09-08-2017** at 03:00 PM and Technical bid will be opened on same day at 03.30 PM.