

Corrigendum

In tender document Tender Enquiry No.24/RT/Nursing/135(I)/2017-RISH (ADMN)

Dated: 27-06-2017

As per schedule, Pre- Bid meeting of "Tender for BLS Mannequin/ Half body CPR for Department of Nursing" was held on 14-06-2017 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No. 24/RT/Nursing/135(I)/2017-RISH (ADMN) has been made.

Point srl no. 02 at page no. 19: -

For: - Should be equipped with a mechanical or electronic monitoring instrument as well as digital sensors showing effectiveness of the resuscitation.

Read as: - Should be equipped with a electronic monitoring instrument as well as digital sensors showing effectiveness of the resuscitation.

Point srl no. 03 at page no. 19: -

For:- It also shows results achieved for insufflations volume and depth of external chest compression, as well as it indicates any stomach inflation and incorrect hand positioning.

Read as: - It also shows results achieved for insufflations volume and depth of external chest compression, and incorrect hand position.

Point srl no. 04 at page no. 19: -

For: It should be incorporates an unique, patented hygienic system under which all trainees get their own face piece and head bag, thereby minimizing any risk of cross infection between the trainees.

Read as: - Cross Infection avoiding method should be available.

Point srl no. 05 at page no. 19: -

For:- The hygienic system also features realistic expiration from the mannequin through the mouth and nose.

Read as:- Deleted.

Point srl no. 06 at page no. 19: -

For:- The rigidity of the chest should be adjustable.

Read as:- The rigidity of the chest should be adjustable to hard, medium and soft Chest.

Consumable

Point srl no. 03 at page no. 19: -

For:- 10 disposable tracheal airways.

Read as:- Deleted.

Point srl no. 05 at page no. 19: -

For:- One resuscitation kit containing AMBU unit.

Read as:- Resuscitation Bag.

Note: Last date of submission of bids in respect of aforesaid tender is hereby extended till **12-08-2017** at 03:00 PM and Technical bid will be opened on same day at 03.30 PM.