

Corrigendum
In tender document
Tender Enquiry No. 24/RT/CTVS/153(II)/2017-Rish(Admn)

Dated: 20-04-2018

As per schedule, Pre- Bid meeting of "Re-tender for Rigid Bronchoscope with accessories for Department of CTVS" was held on 09-04-2018 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry 24/RT/CTVS/153(II)/2017-Rish(Admn)" has been made.

Point Srl no. 21 at page no. 19: -

For: - Micro Scissors, 30 cm length, 4.0 mm shaft diameter- ONE

Read as: - Micro Scissors, length 50 cm and shaft diameter 2.5mm.

Point Srl no. 22 at page no. 19: -

For: - Fiber optic cable 2.5mm Diameter 1.80 meter length- 02

Read as: - Fiber optic cable 2.5-3.5 mm length180 -230 cm.

Point Srl no. 23 at page no. 19: -

For: - Cold light source 250 Watt -1

Read as: - Cold light source Halogen 250W should have built in spare lamp which automatically switch on failure of main bulb.

Point Srl no.24 at page no. 19: -

For: - Mobile Trolley.

Read as:- Mobile Trolley should be same manufacture.