

# **Corrigendum**

## **In tender document Tender Enquiry No.24/Mortuary/160(XI)/2016-RISH (ADMN)**

Dated: 16-12-2016

As per schedule, Pre- Bid meeting of "Autopsy Table with Central Sink and Head rest" for department of Forensic Medicine was held on 05-12-2016 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No. 24/Mortuary/161(XI)/2016-RISH (ADMN) has been made.

### **Specification at page 17 point srl 04:-**

**For:-** Should have the facility of Height adjustment via electro hydraulic mechanism. The minimum height should be 800 mm adjustable upto 1000 mm.

**Read as:-** Should have the facility of Height adjustment via electro hydraulic mechanism. The height should be between 750-850mm and should be adjustable by minimum of 150 mm.

### **Specification at page 17 point srl 14:-**

**For:-** Should have large single compartment sink basin, with approx. dimension of 1 ft X 1 ft X 0.75 ft.

**Read as:-** Should have large single compartment sink basin with approx. dimension of 1 ft X 1 ft X 0.6 ft.

### **Specification at page 17 point srl 28:-**

**For:-** For quality assurance, the firm should have supplied the same make item and model number to any of the six AIIMS / any of the Institutes of National Importance. The bidder should enclose supply order copies along with their technical bid quotation.

**Read as:** For quality assurance, the firm should have **preferably** supplied the same make item and model number to any of the six AIIMS / any of the Institutes of National Importance. The bidder should enclose supply order copies along with their technical bid quotation.

Note: Last date of submission of bids in respect of aforesaid tender is hereby extended till 09-01-2016 at 03:00 PM and Technical bid will be opened on same day at 03.30 PM.