

# **Corrigendum**

## **In tender document Tender Enquiry No.24/ENT/169(I)/2016-RISH (ADMN)**

Dated: 08-02-2017

As per schedule, Pre- Bid meeting of "Tender for Video Esophagoscopy" was held on 16-01-2017 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No. 24/ENT/169(I)/2016-RISH (ADMN) has been made.

### **Specification at last page:-**

1. *"Video Processor should have digital signal processing high resolution color CCD Chip with 2RGB, 2Y/C and 1 BNC Video output connectors for external devices like recording and printing, dedicated keyboard for data entry"* to be replaced by *"Video Processor should have digital signal processing high resolution color CCD Chip with 2 RGB or **other compatible output**, 2 Y/C and 1 BNC Video output connectors for external devices like Recording and Printing, dedicated keyboard for data entry."*
2. *The Video Processor should either have inbuilt Light source with minimum 100 watts Xenon bulb with a standby LED bulb or Bidders can quote Xenon light source separately with same specifications* to be replaced by *The Video Processor should have inbuilt Light source with equivalent to 100 watts Xenon/LED bulb.*

### **Following point is being added:-**

1. Should have facility of recording in the internal buffer memory of processor along with the slot for direct image recording via USB slot a portable device. Should be supplied with a portable data storage device of 1TB minimum and one pen drive of 32 GB.
2. **Trolley: All system should be supplied with stable trolley form the same manufacturer (imported) having space for all components, voltage stabilizer and power sockets that may be required.**

### **Special Conditions:**

- 5 years CMC and 5 years warranty is required including all accessories.
- Should be European CE and USFDA approved.
- All the equipment including all the accessories should be provided with the detailed data sheet along with the original brochure of the quoted model (Photocopies/Print outs are not accepted)
- List of major installations (minimum 5) in last two years should be provided.
- Training for user departments, doctors, and staff should be provided.

Note: Last date of submission of bids in respect of aforesaid tender is hereby extended till 10-03-2017 at 03:00 PM and Technical bid will be opened on same day at 03.30 PM.