

## **Corrigendum-02**

### **In tender document Tender Enquiry No.24/Cardio/174/2016-RISH (ADMN)**

Dated: 04-03-2017

As per schedule, Pre- Bid meeting of "Tender for TMT (Complete Stress System)" was held on 06-02-2017 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No. 24/Cardio/174/2016-RISH (ADMN) has been made.

#### **Point no. 01 at page no. 19-**

**For:** - System should have telemetry module/wireless connectivity (Bluetooth/radiofrequency Based) between Patient and machine to acquire diagnostic quality ECG Data.

**Read as:-** System should have telemetry module/wireless connectivity (Bluetooth /radiofrequency based) between Patient and machine to acquire diagnostic quality ECG Data. (OPTIONAL)

#### **Point no. 02 at page no. 19-**

**For:** - USB Connectivity for high speed data transfer from acquisition device to PC.

**Read as:** - There should be provision for export of soft data from TMT machine to PC (via USB Connectivity) or dedicated machine architecture

#### **Point no. 03 at page no. 19:-**

**For:** - System should be supplied with integrated USFDA approved stress automatic BP measurement device with interface cable to measure automatically the patient NIBP during stress test per the programming done at stress system. Bi-directional data transfer (Compatible with the stress test system) facility should be there.

**Read as:-** - System should be supplied with integrated company specific stress automatic BP Measurement device with interface cable to measure automatically patient's NIBP during stress test per the programming done at stress system. Bi directional data transfer (Compatible with the stress System) facility should be there.

#### **Point no. 05 at page no. 19:-**

**For:-** System should have touch screen of 24" or more display for easy access.

**Read as:-** System should have touch screen of 19 inch or more display for easy access

#### **Point no. 06 at page no. 19:-**

**For:-** should have software inbuilt online support system for controlling system through internet for providing online support.

**Read as:-** should have software inbuilt support system for controlling system through internet for providing online support/ It should have help menu

**In specifications for stress test software at page no. 19: -**

**Point sr. no. 02:-**

**For:-** should display and regularly update ECG 12 Leads 12 medians,1 expanded median, HR, BP, METS, Stage time, test time, protocol name, stage name, speed and grade of treadmill.

**Read as:-** should display and regularly update ECG 12 Lead 12 medians/RAW ECG, HR, BP, METS, Stage time, test time, protocol name, stage name, speed and grade of treadmill.

**Point sr. no. 11 at page no. 20:-**

**For:-** should be able to display ECG in various format like 3 lead+12 median, 6 lead+12 Median, 12 lead+12 median.

**Read as:-** should be able to display ECG in various format like 3 lead+12 median/ RAW ECG, 6 lead+12 Median/ RAW ECG, 12 lead+12 median/ RAW ECG.

**Point no. 13 at page no. 20-**

**For: -** Should have display sensitivities ranging from 0.25 to 8.0 cm/mV"

**Read as:-** Should have display sensitivities ranging from 0.25 and 0.50 cm/mV.

**Point no. 25 at page no. 20-**

**For: -** Should have base line correction (BLC) for stable baseline during test.

**Read as: -** Deleted.

**Point sr. no. 30 at page no. 20:-**

**For:-** Raw data from software should be made available in standard format for further analysis with software like METLAB.

**Read as:-** Raw data from software should be made available in standard format for further analysis with appropriate software (eg. METLAB or equivalent or appropriate software).

**Point sr. no. 39 at page no. 21:-**

**For:-** the display screen must be 24" or more and it should support 1900 X 1200 or 1900 x 1080 resolution and it should display following parameters.

**Read as:-** the display screen must be 19 inch or more and should support 1900\*1200 or 1900\*1080 Resolutions.

**Specification for Treadmill at page no. 21**

**Point srl no. 04:-**

**For:-** Should be USFDA approved.

**Read as:-** Should be USFDA/ European CE approved but this should be company specific & of high quality.

Note: Last date of submission of bids in respect of aforesaid tender is hereby extended till 07-04-2017 at 03:00 PM and Technical bid will be opened on same day at 03.30 PM.