

Corrigendum

In tender document Tender Enquiry No.24/General Surgery/190/2017-RISH (ADMN)

Dated: 12-04-2016

As per schedule, Pre- Bid meeting of "Tender for Ultrasound Machine with Colour Doppler for Department of Department of General Surgery" was held on 20-03-2017 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No. 24/Surgery/190/2017-RISH (ADMN) has been made.

Point srl no. 01 at page no. 19: -

For: - Should be portable; preferably.

Read as: - should be stand-alone system with trolley mounted/ portable.

Point srl no 15 at page no. 19: -

For: - System should be able to capture all frequencies in a single Probe.

Read as: - System should be able to capture multiple frequencies in a single Probe.

Point srl no 16 at page no. 19: -

For: - High dynamic range of minimum 160db or more.

Read as: - High dynamic range of minimum 200dB to 300dB or more.

Point srl no 19 at page no. 19: -: -

For: - Very high frame rate upto 500 frames per second.

Read as: - Very high frame rate upto 500 or more frames per second.

Point srl no. 45 at page no. 20: -

For: - System should be capable of FUSION IMAGING i.e. fusion of Ultrasound image with CT, MRI, PET, CT with Needle tracking system, Auto Registration Motion Artifact Eliminator.

Read as:- System should be capable of FUSION IMAGING i.e. fusion of Ultrasound image with mammography CT, MRI, PET, CT with Needle tracking system, Auto Registration, Motion artifact Eliminator.

Point srl no. 47 at page no. 20: -

For: - Battery life of at least 4 hrs.

Read as: - Should have emergency power backup of minimum 30 minutes or more through in built rechargeable battery or external UPS of appropriate rating.

Following specification points is being added: -

1. Software to be upgradable period of warranty and CMC.
2. On site demonstration is essential.

Note: Last date of submission of bids in respect of aforesaid tender is hereby extended till 03-05-2017 at 03:00 PM and Technical bid will be opened on same day at 03.30 PM.