

## **Corrigendum**

### **In tender document Tender Enquiry No.24/Patho/199(III)/2017-RISH (ADMN)**

Dated: 03-04-2016

As per schedule, Pre- Bid meeting of "Tender for Immunohistochemistry (IHC) for Department of Pathology" was held on 03-04-2017 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No. 24/Patho/199(III)/2016-RISH (ADMN) has been made.

#### **Specification for point srl no. 01 at page no. 19: -**

**For:** - Autostatiner for Immunohistochemistry and In situ Hybridization.

**Read as:** - Autostainer for Immunohistochemistry and In situ Hybridization from Deparrafinization to counter stain with hemataylin/DAPI in single equipment only.

#### **Point srl no. 02 at page no. 19: -**

**For:** - Capacity to handle at least 20-30 slides or more in single run.

**Read as:** - Capacity to handle at least 30 or more slides or more in single run.

#### **Srl no. 03 at page no. 19: -**

**For:** - Completely open system allowing use of primary and secondary reagents from all companies.

**Read as:** - Completely open system allowing use of primary reagents from all companies.

#### **Srl no. 04 at page no. 19: -**

**For:** - Antibody dispensing should be flexible volume of minimum 50µl.

**Read as:** - Deleted.

#### **Srl no. 05 at page no. 19: -**

**For:** - Antibody and micro reagent consumption as per slide should not be more than 100 µl.

**Read as:** - Antibody and micro reagent consumption as per slide should not be more than 150 µl for primary & secondary.....

#### **Srl no. 11 at page no. 19: -**

**For:** - Option of continuous loading of slides and reagent without interruption of runs already in process.

**Read as:** - Deleted.

**Srl no. 18 at page no. 20: -**

**For:** - The system should be fully automated and process IHC, ISH, CISH and FISH tests independently or simultaneously.

**Read as:** - The system should be fully automated and process IHC, ISH, and/ or CISH tests independently or simultaneously.

**Srl no. 25 at page no. 20: -**

**For:** - Modular system with computer and upgradable software. Branded computer with 2 GHZ multicore processor, 6 GB RAM, DVD writer, Original windows 7 Professional or higher, 1 TB HDD, 23" TFT Monitor, Keyboard and Mouse.

**Read as:** - Latest available computer system with compatible software and hardware.

**Srl no. 28 at page no. 21: -**

**For:** - 4 Sets of core staining reagents (for 100 slides) free of cost, for standardization.

**Read as:** - Provide EGFR (Clone EP-22)/ER/PR/ Her2/ new primary antibodies for 100 tests free of cost provide secondary antibodies for 300 tests free of cost for standardization.

**Srl no. 29 at page no. 21: -**

**For:** -Installation process will include satisfactory running of IHC (nuclear, membranous and cytoplasmic staining) and Fish by the company on its own cost. The Installation will not be deemed complete without satisfactory demonstration.

**Read as:** - Installation process will include satisfactory running of IHC (nuclear, membranous and cytoplasmic staining) and ISH by the company on its own cost. The Installation will not be deemed complete without satisfactory demonstration.

**Note:** Last date of submission of bids in respect of aforesaid tender is hereby extended till 21-04-2017 at 03:00 PM and Technical bid will be opened on same day at 03.30 PM