

## **Corrigendum**

### **In tender document Tender Enquiry No.24/ENT/202(III) 2017-RISH (ADMN)**

Dated: 17-04-2016

As per schedule, Pre- Bid meeting of "Tender for Operating Microscope Department of ENT" was held on 10-03-2017 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No. 24/ENT/202(III)/2017-RISH (ADMN) has been made.

#### **Para 2 at page no. 19: -**

**For: -** -Internal motorized fine focusing system, activation by handgrip / foot control panel, continuously adjustable working distance from 200 mm to 500 mm without exchange of objective lens, integrated continuously variable illumination field from 60mm – 15mm or less.

**Read as: -** -Internal motorized fine focusing system, activation by handgrip / foot control panel, continuously adjustable working distance from 200 mm to 450 mm or more without exchange of objective lens. integrated continuously variable illumination field from 60mm – 15mm or less.

#### **Para 8 at page no. 19: -**

**For: - ILLUMINATION SYSTEM:** Coaxial xenon illumination with same Xenon back up with quick-action lamp changer in case of failure of main lamp (350 w).

**Read as: - ILLUMINATION SYSTEM:** Coaxial xenon illumination with same Xenon back up with quick-action lamp changer in case of failure of main lamp (300w-350 w).

**Note:** Last date of submission of bids in respect of aforesaid tender is hereby extended till 02-05-2017 at 03:00 PM and Technical bid will be opened on same day at 03.30 PM.