

Corrigendum

In tender document Tender Enquiry No.24/ENT/202(VI) 2017-RISH (ADMN)

Dated: 21-04-2016

As per schedule, Pre- Bid meeting of "Tender Otologic Micro- Drill System of ENT" was held on 10-03-2017 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No. 24/ENT/202(VI)/2017-RISH (ADMN) has been made.

Point 02 at page no. 19: -

For: - -Shaft is 15⁰ to maximize visualization during drilling up to 16000 rpm.

Read as: - Shaft is 15⁰ to maximize visualization during drilling up to 16000 - 80000 rpm or more.

Note: Last date of submission of bids in respect of aforesaid tender is hereby extended till **08-05-2017** at 03:00 PM and Technical bid will be opened on same day at 03.30 PM.