

Corrigendum

In tender document Tender Enquiry No.24/203(IV)/ 2017-RISH (ADMN)

Dated: 21-04-2016

As per schedule, Pre- Bid meeting of "Tender for Bedside Monitor with Central Monitor for Department of Physiology" was held on 18-04-2017 at 03.00 PM, in the tender office.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No. 24/Common/203(IV)/2017-RISH (ADMN) has been made.

Specification ECG Point 01 at page no. 19: -

For: - Lead type 6 lead.

Read as: - Lead type 12 Lead.

Following point is being added in specification at page no. 19: -

- Display of Perfusion Index.
- Monitor size more than 15 inch.
- Should be able to monitor ECG, HR, ST segment change, QT analysis, SPO2, NIBP, temperature, and respiration,
- Upgradation facility for ETCO2 and IBP in future on site
- European CE or USFDA certificate.
- Inbuilt Battery back up more than 4 hours.

Note: Last date of submission of bids in respect of aforesaid tender is hereby extended till **19-05-2017** at 03:00 PM and Technical bid will be opened on same day at 03.30 PM.