

Corrigendum

In tender document Tender Enquiry No.24/ICU Beds/208/2017-RISH (ADMN)

Dated: 17-04-2016

As per schedule, Pre- Bid meeting of "Tender for Automatic High End Motorized ICU Beds" was held on 13-04-2017 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No. 24/ICU Beds/208/2017-RISH (ADMN) has been made.

1-Point 6 at page no. 2 Line 3 :-

Be less than 50lakh duly certified by chartered accountant
Replaced with- Be less than 2 crore duly certified by chartered accountant

2-Specification line 2 at page no. 19: -

For –Electrically operate trendelenberg & reverse trendelenberg tilt (+-12 degree)

Read as: Electrically operate trendelenberg & reverse trendelenberg tilt(+12 degree) with side tilt & backrest

3- Specification line 5 at page no. 19: -

For: - -Bed should have two piece slit safety side ralling with pendant type controls.

Read as: - Bed should have two piece slit safety side ralling with pendant/embedded /soft drop type controls.

4- Specification line 11 at page no. 19: -

For- Bed should have lockout switches on attendant control pendent /panel

Read as- Bed should have lockout switches on attendant panel & preferably on all control pendent /panel

5- Specification line 14 at page no. 19: -

For- Bed should have built in extension and bed shortening feature with mattress retainer

Read as- Bed should have built in extension 25+/- 5 cm in length with mattress retainer feature

6- Specification line 20 at page no. 20: -

For- Bed should have 125 mm single wheel castor

Read as- Bed should have twin/dual wheel castor greater than 125 mm size branded good quality.

7- Specification line 24 at page no. 20: -

For- The bed should have following dimensions
Height range -32-77cm +/- 5

Read as- The bed should have following dimensions
Height range - 40-77cm +/- 5

8-Specification line 26 at page no. 20: -

For: - Bed should be European CE and as per USFDA safety guidelines.

Read as - Bed should be European CE /USFDA approved but complying with equivalence to USFDA safety standards. A certificate is to be provided for country of origin.

9-Specification line 27 –To be deleted

10- Specification line 32 & 33 –

For- Ingress-IPX4

Ingress Protection (washable beds)-IPX6W

Read As-Ingress protection preferably should be IPX6W but should not be less than IPX5

11- To ADD – Bed should comply with C Arm for X-ray facility with translucent backrest, which may consist of HPL

Note: Last date of submission of bids in respect of aforesaid tender is hereby extended till 05-05-2017 at 03:00 PM and Technical bid will be opened on same day at 03.30 PM.