

# **Corrigendum**

## **In tender document Tender Enquiry No. 24/Pedia Surgery/222(V)/2017-Rish(Admn)**

Dated: 31-07-2017

As per schedule, Pre- Bid meeting of [Tender for Neonatal Ventilators for Department of Pediatric Surgery](#) was held on 16-06-2017 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No. 24/Pedia surgery/222(V)/2017-Rish(Admn) has been made.

### **Part A: Ventilators with HFOV qty.**

#### **Point srl no. 01 at Page no. 20:-**

**For:-** Should be LATEST model, microprocessor controlled able to ventilate patients with body weight from 400 grams to 25 kg

**Read as:-** Should be LATEST model, microprocessor controlled able to ventilate patients with body weight from 400 grams to 40 kg.

#### **Point srl no. 6 at page no. 20: -**

**For:-** Should have US-FDA/European CE certification for neonatal use

**Read as:-** Should have US-FDA and European CE certification.

#### **Point srl no. 10 at page no. 20:-**

**For:-** Invasive ventilation:

**Read as:-** Invasive ventilation: volume and pressure modes both

#### **Point srl no. 64 at page no. 22:-**

**For:-** Reusable & autoclavable Expiratory valve/cassettes – 2 sets, Hinge support arm for the breathing circuit from the same manufacturer – 01 in no., Non-invasive ventilator masks reusable for neonates (3sizes) – 04 set each.

**Read as:-** Reusable & autoclavable Expiratory valve/cassettes – 1 set with free replacement during warranty and CMC, Hinge support arm for the breathing circuit from the same manufacturer – 01 in no., Non-invasive ventilator masks reusable for neonates (3sizes) – 04 set each.

### **Part B: Specification for Ventilators without HFOV - qty 02**

#### **Point srl no. 18 at page no. 23: -**

**For:-** Humidifier – 810: Disposable Chambers & Circuits – 10

**Read as:-** US FDA certified Servo controlled Humidifier : Disposable Chambers & Circuits – 10

#### **Point srl no. 21 at page no. 24:-**

**For:-** Should have US-FDA/European CE certification for neonatal use

**Read as:-** Should have US-FDA and European CE certification.

**Note:** Last date of submission of bids in respect of aforesaid tender is hereby extended till **16-08-2017** at 03:00 PM and Technical bid will be opened on same day at 03.30 PM.