

Corrigendum

In tender document Tender Enquiry No. 24/Pedia Surgery/222(VI)/2017-Rish(Admn)

Dated: 31-07-2017

As per schedule, Pre- Bid meeting of [Tender for Pediatric Ventilators for Department of Pediatric Surgery](#) was held on 16-06-2017 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No. 24/Pedia surgery/222(VI)/2017-Rish(Admn) has been made.

Point srl no. 3 at page no. 20:-

For:- Electromagnetic Compatible, hinged arm holder for holding the circuit.

Read as:- Hinged arm holder for holding the circuit.

Point srl no. 5 at page no. 20:-

For:- Should have inbuilt facility to upgrade with EtcO2.

Read as:- Should have inbuilt facility to measure EtcO2.

Point srl no. 9(i) at page no. 20:-

For:- Tidal volume (Minimum at least 50ml, Maximum up to 2000ml)

Read as:- Tidal volume (Minimum at least 20ml, Maximum up to 2000ml or more)

Point srl no. 9(vii) at page no.20:-

For:- FiO2

Read as:- FiO2 - 21-100%

Following point is being added at page no. 20 after point srl no. 10(xiii):-

xiv) RSBI, LIP & UIP measurements, slow vital capacity and NIF

Following point is being added at page no. 20 after point srl no. 11(ix):-

x). PRVC, APRV, MMV+PSV and VAPS

Point srl no. 13 at page no 20:-

For:- Should have inbuilt exhalation filter.

Read as:- Should have reusable & autoclavable Expiratory valve/cassettes – 1 set with free replacement during warranty and CMC.

Point srl no. 17 at page no. 21:-

For:- Quality Certification: Valid European CE/US FDA.

Read as:- Quality Certification: Valid European CE and US FDA.

Point srl no. 27(V) at page no.21:-

For:- Humidifier – 810: Disposable Chambers & Circuits – 10

Read as:- US FDA certified Servo controlled Humidifier : Disposable Chambers & Circuits – 10

Note: Last date of submission of bids in respect of aforesaid tender is hereby extended till **16-08-2017** at 03:00 PM and Technical bid will be opened on same day at 03.30 PM.