

Corrigendum

In tender document Tender Enquiry No.24/trauma/223/2017-RISH (ADMN)

Dated: 06-07-2017

As per schedule, Pre- Bid meeting of "Tender Battery Operated Sternal Saw and Re-Do Saw with Accessories the Department of Cardiothoracic Surgery" was held on 21-06-2017 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No. 24/trauma/223/2017-RISH (ADMN) has been made.

Point srl no. 01 at page no. 20: -

For: - Hand piece should have titanium body even weight distribution for ideal balance should be light weight and handy least noise making.

Read as: - Handpieces should be of Metallic body, should light weight and handy.

Point srl no. 02 at page no. 20: -

For: - System should be battery operated and have the all electronics part into battery pack.

Read as: - System should be battery operated.

Note: Last date of submission of bids in respect of aforesaid tender is hereby extended till 12-08-2017 at 03:00 PM and Technical bid will be opened on same day at 03.30 PM.