

Corrigendum

In tender document Tender Enquiry No.24/Patho/225(III)/2017-RISH (ADMN)

Dated: 22-06-2017

As per schedule, Pre- Bid meeting of "Tender Trinocular Microscope with Image Projection for the Department of Pathology" was held on 20-06-2017 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No. 24/patho/225(III)/2017-RISH (ADMN) has been made.

Point srl no. 03 at page no. 20: -

For: - Microscope body should have built in LED lamp (at least 12V) without separate lamp housing.

Read as: - Microscope body should have built in LED lamp with powerful illumination without separate lamp housing.

Point srl no. 04 at page no. 20: -

For: - Filters for blue and neutral density.

Read as: - Deleted.

Point srl no 06 at page no. 20: -

For: - 0-80-100 pull out splitter to permit microscope viewing, camera viewing or both together.

Read as:- 0-80-100 or 0-50-100 pull out splitter to permit microscope viewing, camera viewing or both together.

Point srl no. 11 at page no. 20:

For:- Illumination 12V LED pre-centered and prefocussed with built in lamp house which doesnt project out.

Read as:- Illumination LED pre-centered and prefocussed with built in lamp house which doesnt project out.

Point srl no. 13 at page no. 20:

For:- Digital camera- CCD with (12MP) with refresh rate of 25-30 frames per second or better having automatic fast white balance correction for use as projection system.

Read as:- HDNI Connectivity to connect to projector- CCD/CMOS with at least 10N Pat 1920x1080.

Point srl no 25 at page no. 21: -

For:- Microscope and accessories should be fully manufactured by the principal company.

Read as: - Microscope and accessories (including camera and software) should be fully manufactured by the principal company.

Note: Last date of submission of bids in respect of aforesaid tender is hereby extended till **16-08-2017** at 03:00 PM and Technical bid will be opened on same day at 03.30 PM.