

Corrigendum

In tender document Tender Enquiry No. 24/Pulmonary/234(II)/2017-Rish (Admn)

Dated: 18-08-2017

As per schedule, Pre- Bid meeting of "Endo Bronchial Ultrasound System (EBUS) for Department of Pulmonary Medicine" was held on 03-07-2017 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No. 24/Pulmonary/234(II)/2017-Rish (Admn) has been made.

Point srl no. 01 at page no. 20: -

Ultrasound brocho fiber video scope-

- A. Endoscopic Video image should not be marked.
- B. Depth a field 3-50 mm or more.
- C. Channel diameter 2-0mm or more.

Point srl no. (V) at page no. 21,

Light Source - Xenon preferably.

- Inbuilt air pump for pressure regulated leakage testing
- Equipment with NBI/BLI/Full I Scan OE Capability.
- USFDA & European CE approved.

Processor - NBI/BLI/Full I Scan optical enhancement capabilities.

Note: Last date of submission of bids in respect of aforesaid tender is hereby extended till 04-10-2017 at 03:00 PM and Technical bid will be opened on same day at 03.30 PM.