

Corrigendum

In tender document Tender Enquiry No.24/Ortho/237/2017-RISH (ADMN)

Dated: 22-06-2017

As per schedule, Pre- Bid meeting of "Tender Battery Operated Drill for the Department of Orthopedic" was held on 29-06-2017 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No. 24/Orhto/237/2017-RISH (ADMN) has been made.

Point srl no. 01 Modular Two Trigger Handpiece for Drilling & Reaming at Page No. 20:

For: -

- Drill Speed: 1500 rpm
- Ream Speed: 300 rpm
- Drill Torque: 33 in. lbs.
- Ream Torque: 145 in. lbs.
- Weight: Less than 700 gms.

Read as: -

- Drill Speed: 800 to 1500 rpm
- Ream Speed: 200 to 300 rpm
- Drill Torque: 30 to 40 in. lbs.
- Ream Torque: 100 to 145 in. lbs.
- Weight: 1500 to 1750 gms.

Point srl no. 03 Dedicated Saw Hand Piece at page no 20:-

For: -

- Saw Speed: 11,000 cpm
- Arc: 45°
- Weight: Less than 900 gms.

Read as: -

- Saw Speed: 10,000 to 12000 cpm
- Arc: 5°
- Weight: 1500 to 1750 gms.

Point srl no 04 at page no. 20: -

For: - Large NIMH Battery 12 Volt for large bone – 3

Read as:-

Large NIMH/Nicd/Li-ion batteries' 9 to 13 V.

Note: Last date of submission of bids in respect of aforesaid tender is hereby extended till **09-08-2017** at 03:00 PM and Technical bid will be opened on same day at 03.30 PM.