

Corrigendum

In tender document Tender Enquiry No. 24/Nursing/239(III)/2017-Rish (Admn)

Dated: 10-07-2017

As per schedule, Pre- Bid meeting of "Advanced Patient Care Mannequin for College of Nursing." was held on 04-07-2017 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No. 24/Nursing/239(III)/2017-Rish (Admn) has been made.

Following point are being added in specification at page no.20: -

1. It should allow teaching auscultation of normal and pathological hearty bowel sounds and lung sounds.
2. It should be able to teach ECG rhythms with visual display of vital signs like BP, Heartrate, SPO2, etc.
3. It should also be supplied with 03 automatic editable teaching scenarios per manikin to be controlled with wireless device by the nurse instructor.
4. The bidder should provide faculty training for all scenarios preparation, simulation and debriefing.

Note: Last date of submission of bids in respect of aforesaid tender is hereby extended till **12-08-2017** at 03:00 PM and Technical bid will be opened on same day at 03.30 PM.