

Corrigendum

In tender document Tender Enquiry No. 24/Nursing/239(V)/2017-Rish (Admn)

Dated: 10-07-2017

As per schedule, Pre- Bid meeting of "CPR/BLS Mannequin Simulator for College of Nursing" was held on 04-07-2017 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No. 24/Nursing/239(V)/2017-Rish (Admn) has been made.

Point srl no. 03 at page no. 19: -

For: - It also shows results achieved for insufflations volume and depth of external chest compression, as well as it indicates any stomach inflation and incorrect hand positioning.

Read as: - It also *preferably* shows results achieved for insufflations volume and depth of external chest compression, as well as it indicates any stomach inflation and incorrect hand positioning.

Point srl no. 06 at page no. 19: -

For:- The rigidity of the chest should be adjustable.

Read as: - Preferably it should have adjustable rigidity of the chest.

The Following point is being added in specification: -

1. The BLS turbo should be able to connect with wireless tablets, smart phones and/or LCD wired feedback. Providing both student and instructor feedback.

Note: Last date of submission of bids in respect of aforesaid tender is hereby extended till **12-08-2017** at 03:00 PM and Technical bid will be opened on same day at 03.30 PM.