

Corrigendum

In tender document Tender Enquiry No. 24/Urology/254(V) 2017-Rish(Admn)

Dated: 26-09-2017

As per schedule, Pre- Bid meeting of "Adult Cystoscope, Cystolitholapxy Set, Visual Internal urethrectomy (VIU) Set, Bipolar/Saline TURP Set and Bipolar Electrocautery Set for **Department of Urology**" was held on 18-09-2017 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No24/Urology 254(V)/2017-Rish(Admn) has been made.

Point srl no. 2 at page no. 19: -

For: - Automatic Smoke Evacuation facility.

Read as: - Automatic/Compatible Smoke Evacuation facility.

Point srl no. 3 at page no. 19: -

For:- All Instruments should be CE and FDA certified wherever applicable.

Read as:- All Instruments should be European CE and USFDA certified.

Point srl no. 5 at page no. 18: -

For: Cystoscope-Urethroscope sheath, 19/19.5 Fr, with obturator and 2 Luer-lock connectors.

Read as:- Cystoscope-Urethroscope sheath, 19/19.5 Fr, with obturator and 2 Luer-lock connectors.

Point srl no. 7 at page no. 18: -

For:- Cystoscope- Urethroscope sheath, 22/23 Fr, with obturator and 2 Luer-lock connectors.

Read as:- Cystoscope- Urethroscope sheath, 22-23 Fr, with obturator and 2 Luer-lock connectors.

Point srl no. 11 at page no. 18: -

For: - Optical Biopsy forceps, 5Fr and & 7 Fr, each with double action jaws, flexible, length 40cms. - 1pcs

Read as: - Optical Biopsy forceps, 5Fr and & 7 Fr, each with double action jaws, flexible, length 30 cms or more.

Point srl no. 13 at page no. 19: -

For: - Maurmeyer Stone Punch: This set should be compatible with 4mm 12-degree telescope. The set should have the stone punch-working element and should be provided with compatible punch sheath with visual obturator. Sheath should have matt finish for staying of lubrication gel to reduce friction trauma to the patient.

Read as: - Maurmeyer Stone Punch: This set should be compatible with 4mm telescope. The set should have the stone punch-working element and should be provided with compatible punch sheath with visual obturator. Sheath should have matt finish for staying of lubrication gel to reduce friction trauma to the patient.

Point srl no. 16 at page no. 19: -

For:- Laparoscopic Vessel Sealing system should have possibility of vessel sealing upto 7mm, suitable reusable instrument for sealing and cutting simultaneously should be provided (5mm for laparoscopy surgery-2 nos and 10 mm for open surgery-2 nos) (In case quoting Disposable instrument, 3 instruments each type should be quoted).

Read as:- System should have inbuilt vessel sealing capability or should be upgraded to vessel sealing in near future.

Note: Last date of submission of bids in respect of aforesaid tender is hereby extended till **23-10-2017** at 03:00 PM and Technical bid will be opened on **24-10-2017** at 03.30 PM.