

Corrigendum
In tender document
Tender Enquiry No. 24/ICU Beds/283/2017-Rish(Admn)

Dated: 03-01-2018

As per schedule, Pre- Bid meeting of "Tender for ICU Beds" was held on 28-12-2017 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry 24/ICU Beds/283/2017-Rish(Admn)" has been made.

Point Srl no. 28 at page no. 19: -

For: - Bed has polymer molded Safety side railings on both sides.These shall be fitted to the mattress support sections and should be able to raise and lock through spring lock mechanism. When put down, they should undergo and level with the bed frame.

Read as: - Bed has polymer molded 3 split Safety side railings on both sides. These shall be fitted to the mattress support sections and should be able to raise and lock through spring lock mechanism. When put down, they should undergo and level with the bed frame.

Point Srl no.45 at page no. 20: -

For: SS Heavy duty I.V.Pole (2 hooks) (0164).

Read as: - Deleted.

Point Srl no.45 at page no. 20: -

For: - Four Section Mattress with 4" thick PU foam of 40 Density covered with PVC rexine.(0321A).

Read as: - Deleted.

Point Srl no. 45 at page no. 20: -

For: - Utility I.V. Pole with 4 Hooks above angle indicator (0164D).

Read as: - Deleted.

Point Srl no. 45 at page no. 20:-

For: - Telescopic I.V. Pole with 2 hooks above angle indicator (0154B).

Read as: - Deleted.

Point Srl no. 45 at page no. 20: -

For:- Should have oxygen Bottle holder and Urine basket holder

Read as:- Deleted.

Point Srl no. 45 at page no. 20: -

For:- Telescopic I.V. Pole with 4 hooks above angle indicator (0154D).

Read as: - Deleted.

Point Slr no 45 at page no. 20:-

For: - S.S. Lifting pole (0553).

Read as:- Deleted.

Point Slr 45 no at page no. 20:-

For:- Traction pulley attachment for trauma cases. (0187).

Read as:- Deleted.

Point Slr no 45 at page no. 20:-

For: - ICD Bottel Holder (0150).

Read as: - Deleted.

Following points which is mentioned at page no. 20 is being considered as a standard requirement: -

1. Utility I.V. Pole with 2 Hooks above angle indicator (0164B).
2. Urine Bag Holder (0156).
3. Chart Holder Moulded (0158A).
4. Oxygen cylinder cage (Epoxy coated). (0166).
5. Mattress for Bed extension (0122).

Note: - Last date of submission of bids in respect of aforesaid tender is hereby extended till 22-01-2018 technical bid will be opened on next day at 03:30 PM.