Corrigendum-1

In tender document
Tender Enquiry No. 24/Neuro/72(3)/2015-Rish(Admn)
"TENDER FOR HIGH SPEED ELECTRICAL MICRO DRILL OF NEUROSURGERY"

As per schedule Pre- Bid of "High Speed Micro Drill System of Neurosurgery" held on 09-03-2015 at 03.00 PM, in the tender opening room.

The following members attended the meeting:-

Dr. Shobha S Arora, Professor, Chairperson Tender Committee
Dr. Ashok Kumar, Medical Superintendent, Member
Dr. Balram Ji Omar, Assct. Professor, Member
Dr. Yogesh Singh, Asst. Professor, Member
Dr. Rajnish Kumar Arora, Asst Professor, Department of Neurosurgery

The representatives of following firms participated in the Pre-Bid Meeting:-

M/s Johnson & Johnson (P) Ltd
M/s Stryker
M/s India Medtronic Pvt. Ltd.

The following modification (deletions/additions/replacements) for High Speed Electrical Micro Drill of Neurosurgery Tender Enquiry No. F.No 24/Neuro/72(3)/2015-Rish(Admn):-

Last date is submission of bids in respect to the tender of High Speed Electrical Micro drill of Neurosurgery (Enquiry No. 24/Neuro/Eqpt/72(3)2015/Rish(Admn) is hereby extended till 25-03-2015 at 03.00 PM. Technical bid will be open on same day at 03.30 PM in tender opening room.

- ***Existing specification****- Srl No.2*
  For The console should have touch screen display panel.

  Read as- The console should preferably have touch screen display panel.

- ***Existing specification****- Srl No.5*
- *For Should have provision to attach two motor at the same time.*

  Read as – deleted

- ***Existing specification****- Srl No.6*
- *For Should have customizable settings like Acceleration and stopping characteristics for individual motors, oscillation angle.*

  Read as – Should preferably have customizable settings like Acceleration and stopping characteristics for individual motors, oscillation angle.

* ***Existing specification***- *Srl No.8*
- *For The high speed motor should weigh not more than 90 grams & should be smaller than 100 mm in length.*

  Read as- The high speed motor should weigh not more than 120 grams & should be smaller than 120 mm in length.

* ***Existing specification***- *Srl No.15*

- *For System should be supplied with reusable perforator cutter for adult use.*

  Read as- System should be supplied with perforator /cutter for adult and pediatric use.

* ***Existing specification***- *Srl No.19*
- *For A perforated sterilization basket of SS should be supplied. The sterilization basket should have racks to hold the cables, motor and various head pieces. The Sterilisation should be from same manufacturer as that of Console.*

  Read as- A perforated sterilization basket of SS/autoclavable plastic/fiber should be supplied. The sterilization basket should have racks to hold the cables, motor and various head pieces. The Sterilization case should be from same manufacturer as that of Console.

* ***Existing specification***
  Attachments and accessories- *Srl No.2*

- *For Reusable perforator cutter 9/12 - 2 No.*

  Read as- perforator cutter 9/12 or similar size (eg 7/11 or 9/11) - 5 nos

* ***Existing specification***- *Srl No.6*
- *For Straight Hand piece of working length of 40mm - 1 No.*

  Read as- Straight Hand piece of working length of 90-100mm - 1 No.

* ***Existing specification***- *Srl No.7*
- *For Angled Hand piece of working length of 70 mm - 1 No*

  Read as –deleted

* ***Existing specification***- *Srl No.8*
- *For Angled Hand piece of working length of 100 mm - 1 No*

  Read as- Angled Hand piece of working length of 100-110 mm -1 No

* ***Existing specification***- *Srl No.9*
- *For Extra long hand piece for transoral/transphenoidal surgery - 1 no*

  Read as -Extra long hand piece for transoral/transphenoidal surgery of length about 12-13 cm - 1 no.

* ***Existing specification***—Srl No 10
- *For Reusable Tools*

  Read as – Tools

* ***Existing specification****-Srl No 11*
- *For Oscillating Saw - 1 No.*

  Read as - Saggital saw with hand & foot controls-1 no. If system doesnot support both controls
  (i.e. hand and foot) at same time , hand control is preferable.

* ***Existing specification****- Srl No 12*
- *For Oscillating Saw blades - 8 Nos.*

  Read as – Saggital saw blades – one set of different sizes

* ***Existing specification****- Srl No 13*
- *For Motor oil- 6 cans*

  Read as – motor oil, if required- 6 cans

* ***Existing specification****-14,15 & 16*

  Environmental factors
  1. Shall meet IEC-60601-1-2:2001(Or Equivalent BIS) General Requirements of Safety for Electromagnetic Compatibility or should comply with 89/366/EEC; EMC-directive.
  2. The unit shall be capable of being stored continuously in ambient temperature of 0 - 50deg and relative humidity of 15-90%
  3. The unit shall be capable of operating continuously in ambient temperature of 10 -40 deg C and relative humidity of 15-90%

  Read as-

  E. Environmental factors- Should be US FDA/ European CE approved product.

* ***Existing specification-*** on page 13 of tender document- para 18

  For In addition to above –mentions that "tenderer will have to provide complete warranty for all equipments for 2 (two) years &…"

  Read as -It should be warranty of 5 years followed by CMC for 5 (five) years, with an affidavit mentioning the same.

Rest of specifications which are not mentioned here, remain unchanged as in original tender document.