

Corrigendum-1

In tender document

Tender Enquiry No. 24/Neuro/72(6)/2015-Rish(Admn)

"TENDER FOR ULTRASONIC ASPIRATOR OF MICRO NEUROSURGERY"

As per schedule Pre- Bid of "Ultrasonic Aspirator of Micro Neurosurgery" held on 09-03-2015 at 03.00 PM, in the tender opening room.

The following members attended the meeting:-

Dr. Shobha S Arora, Professor, Chairperson Tender Committee
Dr. Ashok Kumar, Medical Superintendent, Member
Dr. Balram Ji Omar, Asst. Professor, Member
Dr. Yogesh Singh, Asst. Professor, Member
Dr. Rajnish Kumar Arora, Asst Professor, Department of Neurosurgery

The representatives of following firms participated in the Pre-Bid Meeting:-

M/s Cure Surgicals
M/s Overseas Associates

The following modification (deletions/additions/replacements) for Ultrasonic Aspirator of Micro of Neurosurgery Tender Enquiry No. F.No 24/Neuro/72(6)/2015-Rish(Admn):-

Last date is submission of bids in respect to the tender of Ultrasonic Aspirator of Micro of Neurosurgery (Enquiry No. 24/Neuro/eqpt/72(6)/2015 Rish(Admn) is hereby extended till 25-03-2015 at 03.00 PM. Technical bid will be open on same day at 03.30 PM in tender opening room.

* **Existing specification**-Page 18 Srl No. 3.4

For The vacuum pump should provide preferable the suction of > 600mm of Hg.

Read As- The vacuum pump should provide preferably the suction of up to 500mm Hg or

more.

* **Existing specification** Page 18 Srl No. 3.7

For It should have integral suction with vacuum pressure of – 0.0 to 0.9 bar in continuous low noise and digital display

Read as-It should have integral suction with vacuum pressure of – 0.0 to 0.6 bar (or better) in continuous Low noise and digital display.

* **Existing specification**- Page 18 Srl No. 3.8

For It should preferably have 1.5 -2.5 liter capacity container of unbreakable material with Level sensor and anti-overflow system.

Read as- It should have 1.5 -2.5 liter capacity container of unbreakable material. The facility of level sensor and anti-overflow system is preferable.

* **Existing specification-** Page 18 Srl No. 3.10

For The hand piece should be available in the following sizes:-

1. Standard Size Hand Pieces- Angled & Straight (1 each)
2. Micro tipped- Angled hand piece. (1 each)
3. Long- Angled hand piece. (1 each).
4. Hand piece angled for bone cutting.

Read As- The hand piece should be available in the following sizes:-

1. short angled Macro hand piece , tip diameter outer -2-2.2mm, inner – 1.4-1.7mm
2. Short angled macro hand piece- tip diameter outer- 3-3.5mm, inner- 2-2.2mm.
3. long angled micro hand piece working length- 100-110mm, weight upto 85g, tip diameter outer 2- 2.3mm, inner 1.5-1.7mm
4. Bone knife/ bone dissector headpiece with disposable titanium cutting tip, length 76mm, cutting tip -7mm×3mm×0.9mm

Addition:-

3.10 (a) - If a 'Single/universal handpiece with changeable tips' is quoted then bidder must quote cost of 60 changeable / disposable tips (standard; n=25, micro precision; n=25 & bone sculpting (short & long angled options; N= 10)

Existing specification- Page 18 Srl No. 3.11

For Should support an endoscopic hand piece.

Read as:- the facility to support an endoscopic hand piece/long hand piece >20 cm is preferable.

* **Existing specification-** Page 18 Srl No. 3.12

For The irrigation pump should be inbuilt in the unit, the irrigation output 0-160ml/min.

Read as- The irrigation pump should be inbuilt in the unit, the irrigation output should be up to 40 ml/min or higher.

* **Existing specifications-** Page 18 Srl No. 4

For System Configuration Accessories, spares and consumables

4.1 ACCESSORIES:

1. Trolley.
2. Assembly kit for aspirator- 1
3. Infusion bottle holder-1
4. Double foot switch-1
5. Cleaning brush for instrument lumen-2
6. Instrument connection cables- 2
7. Suction / irrigation tubing (5meter each), silicon twin tube-20
8. Autoclavable compatible instrument tray.
9. Protective cover-4 pieces.

Read as

4 System Configuration Accessories, spares and consumables

4.1 ACCESSORIES:

1. Trolley.
2. Assembly kit for aspirator- 1
3. Infusion bottle holder/ gallows-1

4. *foot switch-1*
5. *Cleaning brush/wire for instrument lumen-4*
6. *Instrument connection cables- 2*
7. *Suction / irrigation tubing (5meter each), silicon twin tube-20*
8. *Autoclavable compatible instrument tray.*
9. *Protective cover, if required-4 pieces.*

* ***Existing specification-***

For In addition to above – on page 13 of tender document- point 18 mentions that “tenderer will have to provide complete warranty for all equipments for 2 (two) years &...”

Read as - It should be warranty of 5 years followed by CMC for 5 (five) years, with an affidavit mentioning the same.

Rest of specifications which are not mentioned here, remain unchanged as in original tender document.