

Corrigendum-1

In tender document

“TENDER FOR ORTHOPEDICS DEPARTMENT”

After pre-bid meeting on 02-01-2015 at 03.00 PM, the committee recommends for following modification (deletions/additions/replacements) in Tender Enquiry No.F.No 24/Ortho/69(3)/2014-RISH (Admin) related to “Tender for “Arthroscopy System” Orthopedics Department”:-

Last date of submission of bids in respect of Tender for “Arthroscopy System” is hereby extended till **20-01-2015 and Technical bid will be open on same day at 03.00 PM** with following amendments :-

Read as.: Point 2 (ix) is made optional

High definition monitor 26 Inch.

For 1. High Definition monitor, screen 26”, 1080p Full HD wide view monitor with aspect ratio of 16:10

Option for wall mounting and desktop in same unit

Read as(1) High definition LED Monitor, screen 26 inches.

Arthroscope Set

For 3. : Forward oblique Telescope 30°, diameter 2.7mm, length 11 cm

Read as (3) : forward oblique telescope 30 degree dia. 2.7 mm length 10-12 cm

Shaver System

For 1(iii) Shaving RPM 1000-12000, Oscillation mode 500-3000 RPM

Read as 1(iii) Shaving RPM 10000- 12000, Oscillation mode 500-3000 RPM

ACL/PCL

Point No. 1 to 3 Read as

Tendon stripper graduated dia 5-7 mm (all size should be available) length 30 cm

For (7): Femoral Offset Guide, offset 3mm, 4mm, 5.5mm (for transtibial femoral prepatation) - one each

Read as (7) Femoral Offset guide offset 4-6 mm, for transtibial femoral prepatation – 1 each

High definition recoreding system

Add

For 1 : Real time, MPEG 2 HD, 1 or 2 compression engine with full IBP encoding

Read as : Should be full HD

For 2 Video inputs minimum 2 nos. S-Video, 2 nos. Composite, 1 XGA (1024 X 768) and 1 High-

Read as 2 replace (1024 x 768) with 1920 x 1080

Point (3) Definition (1280 x 1024) delete