

Corrigendum-1

In tender document

"TENDER FOR ORTHOPEDICS DEPARTMENT"

After pre-bid meeting on 02-01-2015 at 03.00 PM, the committee recommends for following modification (deletions/additions/replacements) in Tender Enquiry No.F.No 24/Ortho/69(1)/2014-RISH (Admin) related to "Tender for "C-Arm" Orthopedics Department":-

Last date of submission of bids in respect of Tender for C-Arm is hereby **extended till 20-01-2015 and Technical bid will be open on same day at 03.00 PM** with following amendments :-

A. X-ray Generator:-

For (ii) Power output : 2 KW or More

Read as. : Power output min. 8 KW or better

For (vi) Should have facility for continuous fluoroscopy and Pulse fluoroscopy (Pulse rate up to 8 pulse per second)

Read as.: (vi) Should have facility for continuous fluoroscopy and Pulse fluoroscopy (Pulse rate of > 8 pulse per second)