

CORRIGENDUM ANAEROBIC WORKSTATION (DEPARTMENT OF MICROBIOLOGY)

1-(Technical specification) Point 1- Capacity (Petri dish) – Minimum 450

TO BE REPLACED BY- Capacity (Petri dish) – 550 or more

2- Point 2 Airlock capacity minimum 10-12 plate port

TO BE REPLACED BY- Airlock capacity minimum 20-25 plate port

4- Point 10- ANO2 conditions Monitor – should be available

TO BE REPLACED BY- - ANO2 conditions Monitor – should be available preferably with color

5-Point 12 –data logging-Preferable with USB port

TO BE REPLACED BY Data logging-Preferable with USB port with a system capacity to record and save data for more than 1 month

6 –Point no 14 - Microscope- To be provided with LED with image projection with laptop with latest antivirus and windows

TO BE REPLACED BY- To be Provide with Mac based Laptop with latest configuration for storing & downloading data dedicated & specific to anaerobic work.

7- To ADD as Point 31 as –If any time, there is a need of shifting of instrument in warranty period from site of installation to other site, it has to be done by the company engineers /authentic persons free of cost.

8- To ADD as point 32-A training is to be done by the company engineers/technical expert thrice during the warranty period for faculty & technician free of cost.

Note: Last date of submission of bids in respect of aforesaid tender is hereby extended till **05-05-2017** at 03:00 PM and Technical bid will be opened on same day at 03.30 PM.