

Corrigendum

In tender document

Tender Enquiry No.24/Burns & Plastic/167(VII)/2016-RISH (ADMN)

Dated: 16-01-2017

As per schedule, Pre- Bid meeting of "Tender for Micromotor System" for department of Burns & Plastic Surgery was held on 04-01-2017 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No. 24/Burns & plastic/167(VII)/2016-RISH (ADMN) has been made.

Following point is being added after srl no. 04 in General Specification at page no. 19:-

- The Company should have authorized channel partner/ Representative in Rishikesh / Dehradun for immediate support and supply of consumables.

Following point is being added after srl no. d in Main Control Unit at page no. 19:-

- Touch screen for easy operation and change setting
- Can operate two or more simultaneously
- Foot pedal should be dual pedal and both the pedal should run forward and reverse, for non

Micromotor Piece at page no. 19 point srl no. b:-

For: - Scratch proof, titanium.

Read as: - Deleted.

Point Srl no. g:-

For: - The speed must be adjustable from 1000 RPM to 40000 RPM.

Read as: - The speed must be adjustable from 1000 RPM to 40000 RPM or better.

Attachment at page no. 19:-

For: - Point A:-

Read as: - (Separate motor for 3 saws to reduce dependability on one motor. Different saws have different power/torque.

- 2. Speed for Recip saw should be 14000 Cpm or more**
 - Speed for oscillating saw should be 20000 cpm or more**
 - Speed for Sagittal saw should be 25000cpm or more**
- Pin collet and wire collet attachment should be there for Hand & foot Surgeries.

Note: Last date of submission of bids in respect of aforesaid tender is hereby extended till 30-01-2017 at 03:00 PM and Technical bid will be opened on same day at 03.30 PM.