

Corrigendum

In tender document
Tender Enquiry No.24/Common/142(IV)/2016-RISH (ADMN)

Dated: 29-09-2016

As per schedule, Pre- Bid meeting of "Tender For supply of Common Equipments (Syringe Infusion Pump) was held on 27-09-2016 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No. 24/Common/142(iv)/2016-RISH (ADMN) has been made.

Point Srl No. 06 at Page No.2

For: Please attach balance sheet (duly certified by Chartered Accountant) for last three (3) years (Attach copy of annual audited report duly certified by the Chartered Accountant. Minimum turnover which should not be less than 05 crores per annum)

Read as: Please attach balance sheet (duly certified by Chartered Accountant) for last three (3) years (Attach copy of annual audited report duly certified by the Chartered Accountant. Minimum turnover which should not be less than 60 Lakh per annum)

Specification Point srl no. 06 at page no. 19:-

For: Flow range (Rate control mode):

- a. 50ml syringe- 0.1 – 1800.0ml/hr (0.1ml/hr step)
- b. 30ml syringe- 0.1 – 1200.0ml/hr (0.1ml/hr step)
- c. 20ml syringe- 0.1 – 800.0ml/hr (0.1ml/hr step)
- d. 10ml syringe- 0.1 – 400.0ml/hr (0.1ml/hr step)
- e. 05ml syringe- 0.1 – 200.0ml/hr (0.1ml/hr step)

Read as: - Flow range (Rate control mode):

- a. 50ml syringe- 0.1 – 1500.0 or more ml/hr (0.1ml/hr step)
- b. 30ml syringe- 0.1 – 800.0 or more ml/hr (0.1ml/hr step)
- c. 20ml syringe- 0.1 – 600.0 or more ml/hr (0.1ml/hr step)
- d. 10ml syringe- 0.1 – 400.0 ml/hr (0.1ml/hr step)
- e. 05ml syringe- 0.1 – 200.0ml/hr (0.1ml/hr step)

Point srl no. 07 (b) at page no. 19:-

For: Time- 0.1 – 999.9min, 0.1min step.

Read as: - Time- 0.1 – 199.59min, 0.1min step.

Point srl no. 11 at page no. 19:-

For: Three levels of alarm volume: High, Medium and Low.

Read as: - Three or more levels of alarm volume.

Point srl no. 12 at page no. 19:-

For: Three levels of occlusion pressure: High, Medium and Low.

Read as:- Three or more levels of occlusion pressure.

Point srl no. 13 at page no. 19:-

For: Alarms:

c. Air alarm

d. Door open alarm

Read as:- Deleted

Point srl no. 16 at page no. 20:-

For:- Display: Minimum 5 inches or more

Read as:- Display: Minimum 3 inches or more

Point srl no. 18 at page no. 20:-

For: - Should have drug library of more than 1000 drugs.

Read as: - Should have drug library of 100 or more drugs.

Point srl no. 22 at page no. 20:-

For:- Rechargeable, maintenance free battery. Life when fully charged: up to 10 hrs

Read as:-Rechargeable, maintenance free battery. Life when fully charged: 8 hrs or more.

Point srl no. 25 at page no. 20:-

For:-Safety system:

a. Flow clamp- prevents free flow when pump door is opened

b. Air detector- minimizes the risk of air.

Read as:- Deleted.

Note: Last date of submission of bids in respect of aforesaid tender is hereby extended till 21-11-2016 at 03:00 PM and Technical bid will be opened on same day at 03.30 PM.