

Corrigendum

In tender document Tender Enquiry No.24/Patho/225(VI)/2017-RISH (ADMN)

Dated: 22-06-2017

As per schedule, Pre- Bid meeting of "Tender for Decaheaded Microscope for Department of Pathology" was held on 20-06-2017 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No. 24/patho/225(I)/2017-RISH (ADMN) has been made.

Point srl no. 02 at page no. 20: -

For: - Observation tube and Eyepieces:

Main Head: Trinocular observation tube with three-way light distribution (100:0, 20:80, 0:100), have 25MM FOV or more.

Read as: - Observation tube and Eyepieces:

Main Head: Trinocular observation tube with three-way light distribution (100:0, 20:80, 50:50, 0:100), have 25MM FOV or more.

Point srl no 03 at page no. 20: -

For: - Eyepiece: Main observer: 10X magnification with F.O.V 25mm - 2 nos or more. With both side diopter adjustment facilities. LED arrow pointer tricolour, joy stic should be provided by supplier.

Teaching head: for 1+9, 10X magnification with F.O.V minimum **25mm** with both side dioptr adjustment facility.

Read as: - Eyepiece: Main observer: 10X magnification with F.O.V 25mm - 2 nos or more. With both side diopter adjustment facilities. LED arrow pointer tricolour, joy stic for pointer should be provided by supplier.

Teaching head: for 1+9, 10X magnification with F.O.V minimum **22mm** with both side dioptr adjustment facility.

Point srl no 07 at page no. 20: -

For: - Objectives: High performance objective should be suitable for bright field and polarizing application. Plan Fluor 2/2.5, Plan Flour 4X 0.13, Plan Fluor 10X0.30, Plan Fluor 20X0.50, Plan Fluor 40X 0.75, Plan Fluor 100X1.30. Should provide required accessories for polarizer and analyzer for simple polarising.

Read as: - Objectives: High performance objective should be suitable for bright field and polarizing application. Plan Achromat 2/2.5, Plan Achromat 4X 0.13, Plan Achromat 10X0.30, Plan Achromat 20X0.50, Plan Achromat 40X 0.75, Plan Achromat 100X1.30. Should provide required accessories for polarizer and analyzer for simple polarizing.

Point srl no 11 at page no. 19: -: -

For: - Microscope and accessories should be fully manufactured by the principal company. Should provide at least 20 spare bulbs free.

Read as: - Microscope and accessories should be fully manufactured by the principal company.

Note: Last date of submission of bids in respect of aforesaid tender is hereby extended till **12-08-2017** at 03:00 PM and Technical bid will be opened on same day at 03.30 PM.