

Corrigendum

In tender document

Tender Enquiry No. 24/Pediatrics/268/2017-Rish(Admn)

Dated: 15-11-2017

As per schedule, Pre- Bid meeting of "Tender for Multiparameter Monitors for Department of Pediatrics" was held on 09-11-2017 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No 24/Pediatrics/268/2017-Rish(Admn) has been made.

Point srl no. 03 at page no.18: -

For: - Display: colour LED, approx. 10 inch with wide viewing angle, facility for display of atleast 5 waveforms.

Read as: - colour LED, 10 inch or more with touch screen, wide viewing angle, facility for display of atleast 5 waveforms.

Point srl no. 13 at page no.18: -

For: - Trend Display (numerical and graphic) from 24 hrs. Facility for zooming in up to 1 min. The trend data should not be lost on switching off the monitor.

Read as: - Trend Display (numerical and graphic) from 100-240 hrs. Facility for zooming in up to 1 min. The trend data should not be lost on switching off the monitor.

Point srl no. 17 at page no.18: -

For: - Power requirements: 220V / 50 Hz (with adapter) and integral rechargeable batteries (autonomy at least 2 hrs., automatic recharge)

Read as: - Power requirements: 220V / 50 Hz (with adapter) and integral rechargeable batteries (autonomy at least 4 hrs., automatic recharge)