

Corrigendum

In tender document Tender Enquiry No. 24/Urology/269(I) 2017-Rish(Admn)

Dated: 26-11-2017

As per schedule, Pre- Bid meeting of "Tender for C-Arm for Department of Urology." was held on 08-11-2017 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No24/Urology 269(I))/2017-Rish(Admn) has been made.

Point srl no. 32 at page no. 06

For: - The equipment installed should be up for 95% of the total warranty time. If the equipment is down for more than 5% suitable action shall be taken against the supplier including imposition of penalty as deemed fit.

Read as: - The equipment installed should be up for 95% of the total warranty time. If the equipment is down for more than 5% suitable action shall be taken against the supplier including imposition of penalty as deemed fit. Failure on which penalty will be seven days extended warrantee for each lapsed day.

X-Ray Generator 'B'

Point Srl No. 05 at page no. 18: -

For: - mA in fluoroscopy : 0.1 to mA or more in normal fluoroscopy and 12 mA or more in high level fluoro should have facility for continuous fluoroscopy and pulse fluoroscopy (pulse rate upto 8 pulse per second).

Read as: - mA in fluoroscopy : 0.2 to 13 mA should have facility for continuous fluoroscopy and Pulse fluoroscopy (Pulse rate upto 8 pulse per second)

Point Srl No. 06 at page no. 18: -

For: - Should have Digital Spot for high quality single image, 16 mA or mor

Read as: - Should have Digital spot for high quality single image, 13 mA or more

C-Arm mechanism and control panel (digital work station) 'D'

Point Srl No. 02 at page no. 18: -

For: - Motorized vertical movement more than 400mm or more

Read as: - Motorized/counter Balanced manual vertical movements more than 400.

Integrated image processing recording and memory system 'F':-

Point Srl No. 01(a) at page no. 18: -

Image intensifier tube.

For:- Input diameter 9" with Triple field (9/6/4)

Read as: - Input diameter 9" with Triple field (9/6/4) or Dual field (9/6)

Point Srl No. C (1) at page no. 18: -

For:- Medical Grade Monitors (Two nos.)

Read as:- High resolution TFT Monitors (Two Nos.)

Point Srl No. C (2) at page no. 18: -

For: - Min 18 inch or more, black and white, flicker free, high resolution (1280x1024 pixels or better), medical grade flat screen TFT, automatic and manual control of brightness and contrast, mounted on mobile trolley with locking device

Read as:- Min 18 inch or more, black and white, flicker free, high resolution (1280x1024 pixels or better), flat screen TFT, manual control of brightness and contrast, mounted on mobile trolley with locking device.

Digital image processor 'G':-

Point Srl No. 07 at page no. 19: -

For: - Medical imaging software's with ability to store 70,000 images or more in hard disk.

Read as: - Ability to store 70,000 images or more in hard disk.

Point Srl No. 08 at page no. 19: -

For:- DICOM option.

Read as:- DICOM option with print and DICOM dose structured report.

Additional features; H'

For Point srl no. 01 at page no 19: -

For:- Built in UPS to protect & save patient data.

Read as:- UPS to protect & save patient data.

Following point is being added: -

The System should be supplied with UPS of suitable rating.

Free space of C 700 mm or more

Integrated dose measuring chamber to be included.

Note: - Last date of submission of bids in respect of aforesaid tender is hereby extended till 12-12-2017 at 03:00 PM and Technical bid will be opened on next day at 03.30 PM.