

Corrigendum

Tender Enquiry No. F.No.24/Eqpt/36/2013-RIS (Admin)

Corrigendum in Tender for Equipment for Department of Medicine:-

After pre bid meeting on 18-09-2013 at 03.00 PM.

Kindly note the following modifications (deletions/additions/replacements) in the tender for Equipments for Department of Medicine:-

Annexure -I (Page no. 13 of 28)


Point no.36 (1):- Specification of Multipara Monitor


- (iii) **Should be read as:** - Minimum 10 inch wide color TFT LCD
- (iv) **Should be read as:** - Multi- leads ECG, waveforms atleast (7 leads) display
- (viii) **Should be read as:** - Networking capability
- (xi) **Should be read as:** - Compatible with invasive BP, CVP, preferably PCWP & ETCO2 monitoring also.
- (xii) **Should be read as:** - Accessories for measurement of SpO2, TEMP, ECG, PR, RESP, rate and NIBP, invasive BP, CVP should be US FDA/Europe CE.

(Page no. 14 of 28)

Point no.36 (2):- Specification of ECG machine

- (ix) **Should be read as:** - Should have automatic and continuous modes for measurement.


(Dr. Ravindra Kumar)


Dr. Ashok Kumar
Medical Superintendent
AIMS Rishikesh

(Page no. 23 of 28)

Point no.36 (5):- Specification of Defibrillator

2. Operational Requirements

(iv) Should be read as: - Should work on Manual Automated external defibrillator (AED) mode. Manual selection approx 270 J.


3. Technical Specifications

(i) Should be read as: - Should be a Low Energy Biphasic defibrillator monitor with Recorder, having capability to arrest all arrhythmia within maximum energy of 270 joules.

7. Standards, Safety and Training

(i) Should be read as: - Should be US FDA approved product

P. Kumar
(Dr. Rewi Khan)


Dr. Ashok Kumar
Medical Superintendent
AIMS Rishikesh