

Corrigendum
In tender document
Tender Enquiry No.24/General Surgery/133(III)/2016-RISH (ADMN)

Dated: 07-08-2016

As per schedule, Pre- Bid meeting of "Tender for Basic Laparoscopic Set for the Department of General Surgery" was held on 26-08-2016 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No. 24/General Surgery/133(III)/2016-RISH (ADMN) has been made.

Item no. C, Accessories & Instruments (Autoclaveable, dismantable and reusable at page no. 19

For: Length-36Cm, 5mm

Read as: Length should be 33cm or more, 5mm

Item no, C Accessories & Instruments (Autoclaveable, dismantable and reusable at page no. 19):-

For:- Point srl No 17,21,22,23,27,28,29,30,31,32,33,34,35,36,37,38,39, 40,41, 42, 43,44, CUSHERI Liver Retractor, Hassan Cannula.

Read as: These items can be of different make but the company should be reputed & USFDA/European CE approved.

Item no. D Suction and Irrigation Device at page no. 19:-

For:-

- Laparoscopy Instruments Tray Compatible with 3mm, 5mm,10mm instrument.
- Sterilization/Disinfection Tray having sieve tray to lift, Size 27" x 7" x 5" (LxBxD).

Read as:-

These items can be of different make but the company should be reputed & USFDA/European CE approved.

Item no. D Suction and Irrigation Device at page no. 19:-

For:- Formaline Chamber made of Virgin Acrylic 4.5mm thickness, size 26"x8"x8"(LxBxH) with three tray for sterilizing the laparoscope, preferable with three tray.

Read as:-USFDA/European CE is not required for these items.

Note: - Last date of submission of bids in respect of aforesaid tender is hereby extended till 03-10-2016 at 03:00 PM and Technical bid will be opened on same day at 03.30 PM.