

Corrigendum
In tender document
Tender Enquiry No.24/General Surgery/133(I)/2016-RISH (ADMN)

Dated: 10-08-2016

As per schedule, Pre- Bid meeting of "Tender for Choledochoscope Set (with accessories) for the Department of General Surgery" was held on 04-08-2016 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No. 24/General Surgery/133(I)/2016-RISH (ADMN) has been made.

Point Srl No 3 at page no. 19:-

For:

- Distil tip outer Diameter: 15-16 Fr.
- Angle of view: 110 deg.
- Direction of view: 0 deg.
- Deflection of Distal tip: 180 deg up and 100 deg down.
- Working length: 35 -40 cm.
- Instrument channel: 6.5- 7.0 Fr.

Read as:

- Distil tip outer Diameter: 15-16 Fr.
- Angle of view: 110 deg - 120 deg.
- Direction of view: 0 deg.
- Deflection of Distal tip: 180 deg up and 100 deg down.
- Working length: 35 -40 cm.
- Instrument channel: 6.0- 7.0 Fr.

Point Srl No. : 6 at page no. 19:-

For: Three part laparoscopic autoclavable BERCI Grasping Forcep with silicone pads, 360 degree rotational sheath, size 10 mm, length 36 cm, long, double action jaws, with metal handle, can be dismantled with the press of a button.

Read as :Three part laparoscopic autoclavable Grasping Forcep with silicone pads, 360 degree rotational sheath, size 10 mm, length 36 cm long, double action jaws, with metal handle, can be dismantled with the press of a button, either of the same make or other make but should be a European CE/USFDA approved company of repute.

Point Srl No 7 at page no 19:-

For:-BERCI Cholangiography Catheter Guide, for catheters with max. size 2.4mm, with distal angulation. distal 90 angulation downwards.

Read as: Cholangiography Catheter Guide, for catheters with max. size 2.4mm, with distal angulation. distal 90 angulation downwards, either of the same make or other make but should be a European CE/USFDA approved company of repute.

Point Srl No. 8 at page no. 19:-

For: Silicone pads compatible with BERCI forcep insert, package of 10.

Read as : Silicone pads compatible with forcep insert, package of 10.

Point Srl No. 10 at page no. 20:-

For:-BERCI Micro-Knife, pointed, distendable, length 31 cm, size 5 mm.

Read as : Micro-Knife, pointed, distendable, length 31 cm, size 5 mm, either of the same make or other make but should be a European CE/USFDA approved company of repute.

Note: - Last date of submission of bids in respect of aforesaid tender is hereby extended till 13-09-2016 at 03:00 PM and Technical bid will be opened on same day at 03.30 PM.