

Corrigendum

In tender document

Tender Enquiry No.24/Hospital furniture/213/2017-RISH (ADMN)

As per schedule, Pre- Bid meeting of “Tender for Supplying Hospital Furniture” was held on 04-05-2017 at 03.00 PM, in the tender opening room.

After consideration by committee, following modifications (deletions/additions/replacements) for Tender Enquiry No. 24/Hospital Furniture/213/2017-Rish (Admn) has been made.

1. Crash Cart

- **For:** Should have minimum of five drawers with adjustable divider/ Fixed Dividers.
Read as: Should have minimum of Four/Five drawers with adjustable divider/ Fixed Dividers.
- **For:** Size should be :- height:100 to 110cm
Read as: Size should be :- height:100 to 150cm

2. Bed Side Locker

- **For:** Shoes made from hard rubber of uniform texture and chemically inactive to the action of mild acids.
Read as: Castors wheel (4 nos.) on each corner to make it movable and two buffers at rear side for protection.

3. Bed Side Screen (Ward Screen)

- **For:** Bed Side Screen (Ward Screen) Stainless Steel with Single Curtain, Three Section.
Read as: Bed Side Screen (Ward Screen) with single Curtain three section.
- **For:** Middle section should be constructed in two pieces and joined together at the middle
Read as: Deleted

4. Fowler Bed, Four Sections with mattress

- **For:** The Bows should be bolted to the frame on brackets welded to the bows in such a way that bolt or nut should not appear on top surface of the bed frame.
Read as: Deleted
- **For:** Bed head panel with O2 outlet-1 no. and high pressure flexible tube 3 mtr and provision for one side fit in oxygen flow meter, one side fit in o2 outlet (for 10% of bed)
Read as: Deleted

5. Dressing Trolley

- **For:** Two drawers made of SS below upper shelves.
Read as: Deleted

- **For:** Bed head panel with O2 outlet-1 no. and high pressure flexible tube 3 mtr and provision for one side fit in oxygen flow meter, one side fit in o2 outlet (for 10% of bed)

Read as: Deleted

6. Over Bed Table/Food Trolley

- **No Change**

7. Stainless Steel Operation instruments trolley (without bucket ring)

- **No Change**

8. Medicine Trolley

- **For:** Dimension :approx. size 750-760(L)X440-450(W)X840-850(H) mm.

Read as: Dimension : approx. size 750-900(L)X440-550(W)X840-950(H) mm.

9. Patient Examination Table

- **For:** Patient Examination Table

Read as: Patient Examination Table (Standard Height)

10. Examination Couch (Low Height)

- **No Change**

11. Stool SS, Round & Revolving top, adjustable for heights

- **For :**Tubular three legged SS base with legs fitted with thick high quality nylon grommets.

Read as: Tubular three/four legged SS base with legs fitted with thick high quality nylon grommets.

12. Solid Linen Trolley

- **No Change**

13. Wheel Chair

- **For:** Arms:- Two numbers arms duly polished shall be made from seasoned hard hillock wood/Shisham/ Babool and fixed to the tubular arm securely with screw which should not be projected out or injected molded from black polypropylene which should be fixed securely with screws.

Read as: Synthetic padded arm rest

Note: from point number 17 to 30 of technical specification has been deleted as it is the specification of Monitor which has been incorporated due to clerical mistake.

14. Single Foot Step

- **No Change**

15. Foot Step Double

- **No Change**

16. Oxygen Trolley Big-250 CFT Cylinder

- **For:** Framework of the cylinder trolley should be mounted on two 10 cm wheels.

Read as: Framework of the cylinder trolley should be mounted on two 15 cm wheels

- **For:** Mounted on 100/150mm diameter castor for mobility

Read as: Mounted on 150 mm diameter castor for mobility

17. Oxygen Trolley Small-40 CFT Cylinder

- **For :** Should be mounted on 100/150mm diameter castor for mobility.

Read as: Should be mounted on 100mm diameter castor for mobility.

18. Patient Stretcher Trolley

To Add: Patent safety belt should be there

19. Drug Trolley Big with S.S Top

- **No Change**

20. Over Bed Table

- **No Change.**

21. Dressing Trolley

- This item has been deleted as it has been repeated. Item number 5 will be considered in that place.

22. Medicine Trolley

- This item has been deleted as it has been repeated. Item number 8 will be considered in that place.

23. IV Stand

- **No Change**

24. Ward Screen Stainless Steel with Curtain

- This item has been deleted as it has been repeated. Item number 3 will be considered in that place.

25. Plat form /footsteps (Single Step)

- This item has been deleted as it has been repeated. Item number 14 will be considered in that place.

26. Plat form /footsteps (Double Step)

- This item has been deleted as it has been repeated. Item number 15 will be considered in that place.

Note:

Sample of all the hospital furniture asked should be submitted in demonstration before finalization of technical evaluation

Other Terms and Conditions

Point no. 4 of Parameters and Technical Specifications for executing the work (Chapter III, Specifications and allied Technical Details) which states that “Company/manufacturer/firm must have warehouse in the city or within Uttarakhand.(Enclosed address details) the company/manufacturer should be maintaining their office/warehouse in the city/state since last 5 years. Copy of shop and establishment act registration certificate to be enclosed for giving proof of establishment”.

This Point has been deleted.

Point which has been added:

1. If any service issue arises then it should be resolved with 48 hours from the date of complaint lodged with the company otherwise appropriate amount would be deducted from PBG as penalty.
2. If the committee members want to visit manufacturing site then the expense of visit would be borne by the bidder.

Comprehensive warranty of 5 years on all the items is mandatory.

Note: Last date of submission of bids in respect of aforesaid tender is hereby extended till **22-06-2017** at 03:00 PM and Technical bid will be opened on same day at 03.30 PM.