

# **Corrigendum**

## **In tender document Tender Enquiry No.24/Patho/225(II)/2017-RISH (ADMN)**

Dated: 22-06-2017

As per schedule, Pre- Bid meeting of "Tender for Micropipette Sets for Variable Volumes for Department of Pathology" was held on 20-06-2017 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No. 24/patho/225(II)/2017-RISH (ADMN) has been made.

### **Point srl no. 01 at page no. 20: -**

**For: -** Micropipettes should have Adjustable volume, Spring Loaded Tip Cone for connecting tips very tightly

**Read as: -** Micropipettes , adjustable volume, and spring loaded / universal tip cone for connecting tips very tightly of all brands.

### **Point srl no 05 at page no. 20: -**

**For: -** Should have Volume Display: 4 Digits with magnifier with electronic display.

**Read as:-** Should have Volume Display: 3 or 4 Digits volume display.

### **Point srl no 06 at page no. 20: -**

**For:-** Should provide thermal, mechanical and Chemical stability piston and should manufactured from Fortron material/ Non corrosive material.

**Read as:-** Should provide thermal, mechanical and Chemical stability piston and should manufactured from Fortron material or better Non corrosive material.

### **Point srl no 10 at page no. 21: -**

**For:-** Set of micropipette should have one of variable volume range 0.1-2.5  $\mu\text{l}$  ( 2 nos), 0.5 – 10  $\mu\text{l}$ ( 2 nos), 2-20  $\mu\text{l}$  ( 2 nos),10 – 100  $\mu\text{l}$  ( 2 nos),50-200  $\mu\text{l}$ ( 2 nos), 100 – 1,000  $\mu\text{l}$  ( 2 nos), (with appropriate TIPS and autoclavable TIPS box).

**Read as: -** Set of micropipette should have one of variable volume range 0.1 or 0.2-2.0 or 2.5  $\mu\text{l}$  ( 2 nos), 0.5 – 10  $\mu\text{l}$ ( 2 nos), 2-20  $\mu\text{l}$  ( 2 nos),10 – 100  $\mu\text{l}$  ( 2 nos),20 or 50-200  $\mu\text{l}$ ( 2 nos), 100 – 1,000  $\mu\text{l}$  ( 2 nos), (with appropriate TIPS and autoclavable TIPS box).

- 1000 tips per micropipette type.

**Note:** Last date of submission of bids in respect of aforesaid tender is hereby extended till 09-08-2017 at 03:00 PM and Technical bid will be opened on same day at 03.30 PM.