

AIIMS RISHIKESH

INDENT FOR PURCHASE OF STORES

(FORM P-2)

Item: Teaching mannequins for newborn resuscitation (preterm and term) and intubation head

1. Please fill a separate form for each item
2. Please fill completely in triplicate. Incomplete forms and those with illegible writing may not be accepted.

Name of items with full specifications & required accessories	Quantity(in figures and words)	Cost per unit (approx) in foreign currency and Rupees	Total cost (approx)
<p>Teaching mannequins for newborn resuscitation (preterm and term) and intubation head Specifications:</p> <p>Set of mannequins will consist of:</p> <ol style="list-style-type: none"> 4. Advance Preterm Training Manikin (Premature Anne with SimPad Plus) (Laerdal 295-55101) – 1 No. 5. Term Training Mannequin (Baby/Newborn Anne Mannequin) (Laerdal CAT NO. 050000) – 2 Nos. 6. Neonatal/Infant Intubation Trainer Head (Laerdal CAT. NO. 250-00101) – 1 No. <p><u>Advance Preterm Training Manikin</u> (Premature Anne with SimPad Plus)</p> <ul style="list-style-type: none"> • It should be a realistically proportioned 25-week preterm model for resuscitation & critical care. • It should have heart, lung, vocal sounds and cyanosis to facilitate the training in the initiation of proper care and resuscitation of preterm infants. <p>Airway Features</p> <ul style="list-style-type: none"> • Anatomically accurate, realistic airway • Positive Pressure Ventilation • Suctioning • ET tube insertion, Sellick Manoeuvre 	<p>Advance Preterm Training Manikin – 1 (one)</p> <p>Term Training Mannequin – 2 (two)</p> <p>Neonatal/Infant Intubation Trainer Head – 1 (one)</p>	<p>12 lakhs GBP 13025</p>	<p>12 lakhs GBP 13025</p>

Blam

SA

- Right main stem intubation
- OG/NG tube insertion

Breathing Features

- Bilateral and unilateral chest rise and fall with mechanical ventilation
- Cyanosis

Breathing Complications

- Unilateral chest movement (right main stem intubation) with mechanical ventilation

Cardiac

- Realistic Compressions

Vascular Access

- Patent, ready to cut umbilicus with venous and arterial access for bolus or infusion
- Simulated blood flashback upon cannulation of umbilical vein
- Peripheral IV access

Physiological Sounds

- Auscultation of lung sounds during ventilation
- Heart sounds
- Vocal sounds

Training scenarios

It should have following training scenarios written by the American Academy of Pediatrics and supports the Neonatal Resuscitation Program™.

- IX. Positive-Pressure Ventilation and Continuous Positive Airway Pressure (CPAP)
- X. Positive-Pressure Ventilation and Endotracheal Intubation
- XI. Positive-Pressure Ventilation, Endotracheal Intubation, and Chest Compressions
- XII. Positive-Pressure Ventilation, Endotracheal Intubation, Chest Compressions, and Medication
- XIII. CPAP, Oxygen Management, and Orogastric Tube
- XIV. Positive-Pressure Ventilation, CPAP, Intubation, and Surfactant Administration
- XV. Intubation, Chest Compressions and Umbilical Vessel Catheter Placement: Ethics and Care at the End of Life
- XVI. Resuscitation of 25-Week Newborn Twins

Accessories to be supplied with

- Laerdal preterm silicone resuscitator (bag capacity 240 ml) with masks (size 0 & 1) and reservoir bag (600 ml) in compact case - 2

Slam

[Signature]

- Laryngoscope with straight blade (Miller 0 / 00)
- ET Tube 2.5 mm (oral and nasal)
- Peripheral IV - 24 G
- NG/OG tube 6 – 6.5 Fr
- Umbilical Catheter 3.5 Fr Arterial, 5 Fr Venous
- Facemask 42 mm
- Advance model with inbuilt battery - 01
- Basic Training Model - 01
- Birthing lubricant
- Airway Lubricant
- Virtual Blood Concentrate
- Umbilical cords 3 Nos,
- 3M 'Kind removal silicone tape'
- 10ml Syringe
- Cap and Blanket
- Tray for the storage of mannequins
- SimPad plus with 8 preloaded NRP scenarios
- User guide in English including important product information, both hard and soft copies including CDs
- Control Unit with 6 NRP scenarios
- Premature Anne Software License with update cable
- Premature Anne Power Cable / Charger
- AC Adapter, USB Cable, wrist strap, Carrying Case

Training and necessary documents

- The bidder should provide training and necessary documents including suggested learning objectives, scenario set up and preparation, equipment checklist, role play information, proposed correct interventions, suggestive guided reflective questions and a section with clinical points to consider during simulation and debriefing along with the interventions that are logged by the instructor during simulation based training shall be available with each above mentioned scenarios.
- The service provider should have the necessary equipment's recommended by the manufacturer for installation and to carry out preventive maintenance test as per guidelines provided in the service/maintenance manual.

Blame
lp

Warranty

- One year against any manufacturing defect

Term Training Mannequin

(Baby/Newborn Anne Mannequin)

- Specifically designed with manikin features that are clinically relevant to acquire the skills in training for neonatal life support
- Educational effective allowing specific neonatal resuscitation skills to be taught individually or in combination with other skills.
- Mannequin represents a full term (40 week), 50th percentile newborn female, measuring 21 inches and weighing 7lbs.
- The airway allows for training in all aspects of newborn airway management, including the use of positive-pressure airway devices, and the placement of ET tubes and LMAs.
- The torso includes functionality to relieve a tension pneumothorax via needle decompression.
- The patent umbilicus has a manually generated pulse and can be assessed, cut and can be catheterized for IV access.
- Features IO access in both legs.

Glamm

Airway Features

h

- Positioning the newborn to simulate opening the airway via head tilt, chin lift or jaw thrust
- Positive Pressure Ventilation (BVM, T-Piece resuscitator, or anesthesia bag)
- ET tube intubation
- LMA insertion
- Orogastric tube insertion
- Stomach distension (when ET is misplaced)
- Suctioning (of the nares, nasopharynx, oropharynx, esophagus and the lungs via an ET tube) - Fluids should not be introduced into the airway
- Meconium module for suction removal

Breathing features

- Bilateral and unilateral (with mainstem intubation) chest rise and fall with mechanical ventilation
- Pneumothorax - Needle thoracentesis left mid axillary (pneumothorax)

Cardiac features

- Manual chest compression at

appropriate depth (1/3 AP) and force
Circulation Features

- Manual umbilical pulse

Vascular Access

- Umbilical Vein/ Artery access via patent umbilicus
- IO access in left and right lower leg, tibial tuberosity and medial malleolus

Other Features

- Full articulation

Accessories to be supplied with

- Laerdal pediatric silicone resuscitator (500 ml bag capacity) with term masks (size 0, 1) and reservoir bag (600 ml) in compact case – 2 Nos
- Soft case, lubricant

Training and necessary documents

- The service provider should have the necessary equipment's recommended by the manufacturer for installation and to carry out preventive maintenance test as per guidelines provided in the service/maintenance manual.

Warranty

- One year against any manufacturing defect

Laerdal Neonatal/infant Intubation

Trainer

- Should allow teaching of intubation skills on the newborn baby
- The mannequin should have realistic anatomy of a newborn baby
- Should allow teaching intubation (both oral and nasal)
- Should have facility for Bag-Valve Mask ventilation
- Correct tube placement can be checked by practical inflation tests.

Accessories to be supplied with

- Lung set – neonatal intubation trainer
- Base & cover assembly
- Mannequin lubricant
- Neonatal (small) laryngoscope handles with Miller straight blades (0, 1) – 2 Nos

Training and necessary documents

- The service provider should have the necessary equipment's recommended by the manufacturer for installation and to carry out preventive maintenance test as per

Slamm
h

guidelines provided in the service/maintenance manual. Warranty <ul style="list-style-type: none"> One year against any manufacturing defect 			
---	--	--	--

3. For equipment, please provide the following information

Detailed description of the actual use of the equipment: Resuscitation mannequins are essential equipments necessary for day to day teaching and training of students, nursing officers and residents.

Is the equipment to be used for patient care or research: Yes

If both, state % of time to be used for patient care: 50% of time to be used for research

Is this/ similar equipment already available in the department? No

When purchased? Cost at that time: Present functional status:

Tests/ procedures done on this equipment in last year:

Revenue generated by this equipment in last year:

If yes, what is the justification for this purchase?

Is this/similar equipment available in any other department in the Institute?

If yes, what is the justification for this purchase?

Slamm

4. For Consumables, please provide following information:

Description of stocks available

When was it last purchased? In what quantity? Cost;

Source

Test/ procedures done in this period:

Revenue generated in this period:

Average annual consumption

Shelf life

Period for which this purchase will last Number of tests likely to be done with this quantity:

Sp

5. For furniture, please provide the following information:

Exact location and use

Existing furniture at that place

Justification for this purchase

Possible sources (name all sources you know) from where item may be obtained (name, address, phone no, fax no, email, etc of contact person)