

**F.No.24/Pedia/61(4), (5),(9), (10) &(11)/2014-RISH(ADMN)/**  
**ALL INDIA INSTITUTE OF MEDICAL SCIENCES, RISHIKESH**

**VIRBHADRA MARG,**  
**RISHIKESH -249201.**

Dated: - 19<sup>th</sup> Dec, 2014

**Subject- Report of Technical Committee against Tender No. 24/Pedia/61(4),**  
**(5), (9), (10) & (11)“**

Sir,

The Technical Committee of AIIMS Rishikesh has evaluated the Tender Documents [No. **24/Pedia/61(4), (5), (9), (10) & (11)/2014-RISH(ADMN)**], relating to equipment of “Pediatrics Department” and submitted their report to tender committee.

As per the report of the Technical Committee, you are, requested to demonstrate the under mentioned equipment to technical committee at Tender opening room AIIMS, Rishikesh between **26-12-2014 to 31-12-2014 at 09.00 AM to 01.00 PM** otherwise your bid will be declared technically disqualified.

The list of firms with items which is to be demonstrate mentioned against each:-

Srl No	Name of the firm	Items
01	M/s Global Medical System, Delhi	Qualified for “ <b>Benchtop Billirubinometer</b> ” subject to onsite demonstration & clarification on validity of CE Certificate.
		Disqualified for “ <b>Fiberoptic Phototherapy</b> ” due to Major deviation Fiberoptic pad surface area : 260 sq cm (<500 sqcm)
02	M/s CL Micromed Pvt Ltd, Delhi	Qualified for “ <b>Benchtop Billirubinometer</b> ” subject to onsite demonstration.
03	M/s PV Medipoint, Haryana	Qualified for “ <b>Fiberoptic Phototherapy</b> ” subject to onsite demonstration.
04	M/s Wipro GE Healthcare, Gurgaon	Qualified for “ <b>Fiberoptic Phototherapy</b> ” subject to onsite demonstration.
05	M/s Vishal Surgical Equipment Co. Delhi	Qualified for “ <b>Resuscitation Training Mannequins- Neonatal</b> ”.
		Qualified for “ <b>Resuscitation Training Mannequins- Pediatrics</b> ”.
		Qualified for “ <b>Intubation Head</b> ”.
06	M/s Hansraj Nayar (Mediquip), Amritsar	Qualified for “ <b>Resuscitation Training Mannequins- Neonatal</b> ”. Subject to ISO/equivalent certification.
		Qualified for “ <b>Resuscitation Training Mannequins- Pediatrics</b> ”. Subject to production of ISO/equivalent certificates.
		Qualified for “ <b>Intubation Head</b> ”.