

Corrigendum-1

In tender document

“TENDER FOR PEDIATRICS DEPARTMENT”

After pre-bid meeting on 08-10-2014 at 04.00 PM, the committee recommends for following modification (deletions/additions/replacements) in Tender Enquiry No.F.No 24/Pedia/61 (5) & (11)/2014-RISH (Admin) related to “Tender for Pediatrics Department”:-

- (c) Tender Enquiry No.24/Pedia/61(5)/2014-Rish(Admn): Fibroptic Phototherapy
- Pg 18 (description): following points should be read as:
 - I. Halogen lamps/LED light optic assembly
 - II. Should emit radiation of wavelength 425-475 nm. It should be able to provide average spectral irradiance between 30 and 45 μ W/cm²/nm
 - III. The optic fibre pad should be flexible, well insulated, waterproof and easy to clean. It should provide exposure surface area of at least 500 cm²
 - Pg 18 (System Configuration Accessories, spares and consumables): There should be free of cost replacement of light source (halogen/LED) for upto 20,000 hours usage of the unit.
 - Pg 18 (Standards, Safety and Training): Electrical safety conforms to standards for electrical safety IEC-60601 / IS-13450/13485 or equivalent
- (d) Tender Enquiry No.24/Pedia/61(11)/2014-Rish(Admn): Intubation head
- 18 (Technical specifications): Neonate Intubation Manikin: It is used for training of endotracheal intubation in resuscitation teaching activities.
Features: 1. Realistic anatomy of neonatal head and neck with attached expandable lungs