

ALL INDIA INSTITUTE OF MEDICAL SCIENCE
RISHIKESH
Tender for Syringe Infusion Pumps
Tender No. 24/Common/142(IV)/2016-Rish (Admn)

Date-08/03/2017

Revised Report of the Store Purchase Committee

The Tender Opening Committee opened the tender for Syringe Infusion Pump against Tender Enquiry No. 24/Common/142(IV)/2016-Rish(Admn) in front of participating bidders on 09/12/2016 and found that 05 tenders received from following firms.:-

S/No	Firm Name
A 1	M/s B Braun Medical (india) Pvt. Ltd., Delhi
2	M/s Meditron, Dehradun
3	M/s POCT Services, Lucknow
A 4	M/s Relief Surgical & Medical Equipment, Dehradun
5	M/s Surya Dev Healthcare Pvt. Ltd., New Delhi

After administrative evaluation of all the tenders received, Store Purchase Committee found that both the firms fulfil eligibility criteria for qualification of technical bids and thus declared qualified for further technical evaluation and demonstration.

Since the Infusion Pumps are required by most of the departments, a Technical Evaluation Committee has been constituted with the HoD of the indenter departments including the HoD's of the aligned departments.

The Technical Evaluation Committee evaluated the technical bids submitted its report to the tender office. As per the report the below firms are declared disqualified because of the reasons mentioned against them.

S/No	Firm Name	Remarks
1	M/s POCT Services, Lucknow	Disqualified Working modes are 3 instead of 5 (demanded in tender) Max infusion rate for 10 ml syringe is 200 ml (instead of demanded 400 ml/hr)
2	M/s Surya Dev Healthcare Pvt. Ltd., New Delhi	Disqualified Max flow rates are less as compared to tender specs Battery life 7 hours (instead of demanded 8hrs or more)

The Technical Evaluation Committee conducted the demonstration and used the Syringe Infusion Pumps submitted by the qualified bidders for one week and submitted its report to the tender office. As per the report the Syringe Infusion Pumps submitted by M/S Meditron is not user friendly and difficult to use in emergency situations quickly. Hence the firm was declared disqualified.

S/No	Firm Name	Remarks
1	M/s B Braun Medical (india) Pvt. Ltd., Delhi	Qualified
2	M/s Relief Surgical & Medical Equipment, Dehradun	Qualified

In the meantime a representation has been received from M/S Meditron, Dehradun that the disqualification was not correct and the committee should re-consider the decision.

The Technical Evaluation Committee accepted the representation and a demonstration was conducted again by all the committee members on 07-03-2017 for M/S Meditron in the Cardiology Ward. In the demonstration it was found that the quoted model by the firm was capable enough to handle emergency situation and can be declared qualified for opening of financial bid.

Financial bid of aforesaid 3 eligible firms will be open on **20/04/2017 at 03.30 PM** in the Tender Office, AIIMS Rishikesh. All successful bidders are requested to participate in the opening of Financial Bid.