

**Corrigendum-2**  
**In tender document**  
**Tender Enquiry No.24/pedia surgery/181(III)/2016-RISH (ADMN)**

Dated: 29-03-2017

After issuing corrigendum 1 it is noted by SPC/Department of Pediatric Surgery that modification made in corrigendum 1 at page 1 may be read as following:-

**Medium PCNL Set: page no. 19:-**

**Point srl no. I (ii):-**

**For:-** Outer Diameter- 20Fr

**Read as:-** Outer Diameter-18-22Fr

**Point srl no. I (iii):-**

**For:-** Viewing direction- 6-10 degree.

**Read as:-** Viewing direction-10-30 degree.

**Point srl no. I (v):-**

**For:-** Working length- 220mms

**Read as:-** Working length- 210-230mms.

**MINI PCNL Set: page no. 19:-**

**Point srl no. I (ii):-**

**For:-** Outer Diameter- 12Fr.

**Read as:-** Outer Diameter- 10-12Fr.

**Point srl no. I (iii):-**

**For:-** Viewing direction- 6-10 degree.

**Read as:-** Viewing direction- 6-30 degree.

**Point srl no. I (iv):-**

**For:-** Working channel- 6-10 Fr.

**Read as:-** Working channel- 6-8Fr.

**Page no. 20:-**

**For:-** *ULTRA mini PCNL Set.*

**Read as:-** Deleted.

**URS Sets at page no. 20:-**

**Rigid Cystoscope:**

**Point srl 01:-**

**For:-** Continuous flow system.

**Read as:-** Deleted

**Note: Rest amendment will remain same as issued corrigendum 1 page no.2:**