

Corrigendum

In tender document Tender Enquiry No.24/Patho/143(II)/2016-RISH (ADMN)

Dated: 03-09-2016

As per schedule, Pre- Bid meeting of "Tender for Phase Contrast Microscope for Department of Pathology" was held on 29-08-2016 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No. 24/Patho/141(II)/2016-RISH (ADMN) has been made.

Point Srl No. 01 at page no. 17:-

For: Optical System: Infinity Corrected Optical System for Better Image contrast and Future Up-gradation.

Read as: Optical System: Infinity Corrected Optical System for Better Image contrast and Future Up-gradation to DIC, Fluorescence, Polarizing, dark field etc.

Point Srl No. 02 at page no. 17:-

For: Stand: Ergonomic design Microscope stand for brightfield & Phase contrast Microscopy with coarse and fine focusing knobs. Suitable High Contrast long Life LED or Suitable Halogen Transmitted Light, Daylight filter, Cover and Power cord. External power supply unit for transmitted light so that no heat transfer to sample.

Read as: - Stand: Ergonomic design Microscope stand for brightfield & Phase contrast Microscopy with 3 step coarse, medium and fine focusing knobs. Suitable High Contrast long Life LED Transmitted Light, Daylight filter, Cover and Power cord. External power supply unit for transmitted light so that no heat transfer to sample.

Point Srl No. 03 at page no. 17:-

For: Nosepiece: 6 position revolving nosepiece to accommodate 6 objective lens of different magnifications at a time.

Read as:- Nosepiece: 6 or more position revolving nosepiece to accommodate 6 or more objective lens of different magnifications at a time.

Point Srl No. 04 at page no. 17:-

For:- Mechanical Stage: XY mechanical stage with holders to adapt glass slides, etc.

Read as:- Mechanical Stage: XY mechanical stage with holders to adapt glass slides, etc. and should be usable by both right and left hand users

Point Srl No. 05 at page no. 17:-

For:- Condenser: Suitable condenser for Brightfield and Phase Contrast.

Read as: Condenser: 5 or more position universal condenser with turret for Brightfield and Phase Contrast.

P.T.O.

Point Srl No. 08 at page no. 17:-

For: Digital Camera: Digital colour camera capable of handling very low light, darkfield or DIC images with 2/3" high-density CCD/CMOS chip, approximately 12.7 million pixel resolution (2200 TB lines), 15 f/p/s with full screen size, cooling 10 degree celsius below ambient, 12 bit digitization, exposure time 1/16000 to 60 seconds, dynamic range 2000:1, USB port for attaching camera onto desktop/laptop through single wire.

Read as:- Digital Camera: Digital colour camera capable of handling very low light, darkfield or DIC images with 2/3" or larger high-density cooled CCD/CMOS chip, approximately 10-16 million pixel resolution (2200 TB lines), 15-30 f/p/s with full screen size, , 12 bit digitization, exposure time upto 500milli seconds, , USB port for attaching camera onto desktop/laptop through single wire.

Point Srl No. 09 at page no. 17:-

For:- Image Analysis Software: The software should have the following features:-

- i. Multi-channel & Time-lapse FL imaging, Movie Recording
- ii. Intensity measurements, line profile, spatial measurements such as length, width, area, perimeter, etc.
- iii. Microscope & camera should be controlled by the microscope manufacturer's software for better compatibility.

Read as:- Image Analysis Software: The software should have the following features:-

- i Multi-channel & Time-lapse FL imaging, Movie Recording
- ii Intensity measurements, line profile, spatial measurements such as length, width, area, perimeter, etc.
- iii Microscope, camera and soft ware should be supplied by the same manufacturer for better compatibility.

Note: Last date of submission of bids in respect of aforesaid tender is hereby extended till 04-10-2016 at 03:00 PM and Technical bid will be opened on same day at 03.30 PM.