

## **Corrigendum**

### **In tender document Tender Enquiry No.24/Ophtho/136(VI)/2016-RISH (ADMN)**

Dated: 27-08-2016

As per schedule, Pre- Bid meeting of "Tender for Phacoemulsification System or Department of Ophthalmology" was held on 22-08-2016 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No. 24/Ophtho 136(VI)/2016-RISH (ADMN) has been made.

#### **Ultrasound Point srl no. 01 at Page no. 19:**

**For:** Should have the ability to drive high performance four crystal handpiece: 40kHz, piezoelectric, slim, and lightweight and autoclavable.

**Read as:** Should have the ability to drive high performance four crystal handpiece: 38 kHz & above, piezoelectric, slim, and lightweight and autoclavable.

#### **Ultrasound Point srl no. 04 at Page no. 19:**

**For:** The ultrasound and neosonix handpiece should be compatible with tips like standard 30 degrees and 45 degrees, microtip, and kelman, flared and aspiration bypass tips.

**Read as:** The ultrasound and neosonix handpiece should be compatible with tips like standard 30 degrees and 45 degrees, microtip, bent & straight aspiration bypass tips.

#### **Ultrasound Point srl no. 08 at Page no. 19:**

**For:** Should have peristaltic system and the facility to use vacuum level 600 650 mmhg and aspiration flow rate 50-60cc/min.

**Read as:** - Vacuum level should be 600mmhg and above.

#### **Note:-**

1. On site demonstration may be required.
2. Last date of submission of bids in respect of aforesaid tender is hereby extended till 13-09-2016 at 03:00 PM and Technical bid will be opened on same day at 03.30 PM.