

Corrigendum

In tender document Tender Enquiry No.24/Pathology/143(VI)/2016-RISH (ADMN)

Dated: 03-09-2016

As per schedule, Pre- Bid meeting of "Tender for Automated Cover Slipper Department of Pathology" was held on 29-08-2016 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No. 24/Patho/141(VI)/2016-RISH (ADMN) has been made.

Point Srl No. 04 at page no. 17:-

For: Should be capable of being integrated with automated stainers.

Read as: It is desirable if independent Automated cover Slipper units are also compatible with automated slide stainers.

Note: Last date of submission of bids in respect of aforesaid tender is hereby extended till 04-10-2016 at 03:00 PM and Technical bid will be opened on same day at 03.30 PM.