

Corrigendum

In tender document Tender Enquiry No.24/Pathology/143(I)/2016-RISH (ADMN)

Dated: 03-09-2016

As per schedule, Pre- Bid meeting of "Tender for Automated Slide Stainer for Department of Pathology" was held on 29-08-2016 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No. 24/Patho/141(I)/2016-RISH (ADMN) has been made.

Point Srl No. 02 at page no. 17:-

For: High throughput robotic stainer for multiple staining applications and should run up to 12 racks/ baskets in parallel.

Read as: High throughput robotic stainer for multiple staining applications and should run up to AT LEAST 12 racks/ baskets in parallel.

Point Srl No. 06 at page no. 17:-

For: The equipment should have at least 30-35 reagent stations and 6 wash stations of 450ml capacity.

Read as: - The equipment should have at least 25-35 reagent stations and 6 wash stations of capacity 300 ml or above.

Point Srl No. 09 at page no. 17:-

For: Option for slide drying should be available.

Read as: INBUILT HEATING STATION for slide drying should be available.

Point Srl No. 18 at page no. 17:-

For: The equipment should be compatible for integration with any automated cover slipper

Read as: It is desirable if the equipment will be compatible with any independent automated cover slipper unit.

Note: Last date of submission of bids in respect of aforesaid tender is hereby extended till 04- 10-2016 at 03:00 PM and Technical bid will be opened on same day at 03.30 PM.