

Corrigendum

In tender document Tender Enquiry No.24/Patho/143(VIII)/2016-RISH (ADMN)

Dated: 03-09-2016

As per schedule, Pre- Bid meeting of "Tender for Fully Motorised Microtome for Department of Pathology" was held on 29-08-2016 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No. 24/Patho/141(VIII)/2016-RISH (ADMN) has been made.

Point Srl No. 04 at page no. 17:-

For: Course feed: Motorized coarse feed to include 300 micron/sec and 900 micron/sec.

Read as: Course feed: Motorized coarse feed OF VARIABLE SPEED SHOULD BE PRESENT.

Point Srl No. 07 at page no. 17:-

For: Trimming section thickness: 1 to 300 microns or more.

Read as: - Trimming section thickness: 5 OR LESS to 300 microns or more.

Point Srl No. 08 at page no. 17:-

For: Specimen retraction: 5 to 100 micron in 5 micron increments can be turned off (manual operation).

Read as: - Specimen retraction: can be turned OFF and On with different increments

Note: Last date of submission of bids in respect of aforesaid tender is hereby extended till 04-10-2016 at 03:00 PM and Technical bid will be opened on same day at 03.30 PM.