

Corrigendum

In tender document of

“Re-tender for Non-Invasive Ventilator (Bipap Machine)”

As per schedule Pre- Bid of “Re-tender for Non Invasive Ventilator (Bipap Machine)” held on 20-10-2015 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following (deletions/ additions/ replacements) for Tender Enquiry No. F. No 24/Re-tender/pulmonary/Eqpt/29(II-1/2015/Rish (Admn) has been done.

Specifications of NIV/Bi-Level Ventilator suitable for hospital use

Annexure -1:

Page No. 19 Para 2:

For: Modes of Ventilation: S/T (Spontaneous/Timed), PAC (Pressure Assisted Control), CPAP (Continuous Positive Airway Pressure), S (Spontaneous), T (Timed).

Read as: Modes of Ventilator S/T (Spontaneous/Times), PAC (Pressure Assisted Control), CPAP (Continuous Positive Airway Pressure), S (Spontaneous), T (timed), PCV, VCV or More.

Page No. 19, Para 12:

For:

Weight: Less than 3 Kgs.

Read As:

Weight: Less than 5.5 Kgs.

Page No. 20, Para 12:

For:

Air Filter: Electrostatic fibre mesh.

Read As:

Air Filter: Electrostatic fibre mesh or equivalent.

Additional point for page No. 20.

Tidal Volume: 100ml to 2500ml or more.